RECEIVED
By Carol Day at 12:19 pm, Mar 22, 2016

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER	LOCATION OF INSTRUMENT	DATE OF INSPECTION	TIME OF INSPECTION
80-005855	CARTHAGE POLICE	03/18/2016	03:11

CALIBRATION CHECK RESULTS

Test	g/210L	Time
Air Blank	0.000	03:13
Cal Check	0.099	03:14
Air Blank	0.000	03:14
Cal Check	0.099	03:15
Air Blank	0.000	03:15
Cal Check	0.099	03:16
Air Blank	0.000	03:16

Pass

CALIBRATION CHECK SUMMARY

STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
DRY	AG408501	03/26/2016
SIM TEMPERATURE	**SIM SERIAL NUMBER**	**SIM CERTIFICATE EXPIRATION**
N/A	N/A	N/A
STANDARD VALUE	**STANDARD SUPPLIER**	
0.100	INTOXIMETERS	

CALIBRATION CHECK RESULT 1: 0.099

CALIBRATION CHECK RESULT 2: 0.099

CALIBRATION CHECK RESULT 3: 0.099

MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)
1.0%	0.000

DIAGNOSTIC TEST RESULTS

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temperature Regulation Test	Pass

Pass

RFI TEST RESULTS

Test	g/210L	Time
Air Blank	0.000	03:17
Subject Test	RFI*	03:17
Air Blank	0.000	03:17

*RFI Detect

Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT

REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
1	1	2	3	1	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

MARCH 2016

INSPECTING OFFICER

SIGNATURE	PRINT NAME	
	BUTLER,JUSTIN	
TYPE II PERMIT NUMBER	**EXPIRATION DATE**	**TELEPHONE NUMBER**
240427	12/12/2016	4172377200

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 26-Mar-2014

Lot # AG408501

Exp. Date	Cyl. Type	Component	Certified Concentration
26-Mar-2016	108	Ethanol	0.100 ± 2% BrAC (260 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

Serial No.	Concentration	Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2014.03.27 10:08:05 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Analyst: Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

JUSTIN BUTLER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/12/2014

NUMBER 240427

EXPIRES 12/12/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Gail Vasterling ,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BUTLER, JUSTIN
Permit No 240427
Date Issued 12/12/2014 **Date Expires** 12/12/2016