



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

**RECEIVED**

REPORT #2

By Carol Day at 10:27 am, Mar 01, 2016

Complete this report in duplicate at the time of the regular monthly preventive maintenance or when the instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-006267	LOCATION OF INSTRUMENT GRAIN VALLEY PD	DATE OF INSPECTION 02/29/2016	TIME OF INSPECTION 13:13
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # AG502603	STANDARD EXPIRATION DATE 01/26/2017
Air Blank	0.000	13:14	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Cal Check	0.081	13:14	STANDARD VALUE 0.080	STANDARD SUPPLIER INTOXIMETERS	
Air Blank	0.000	13:15	CALIBRATION CHECK RESULT 1 0.081		
Cal Check	0.082	13:15	CALIBRATION CHECK RESULT 2 0.082		
Air Blank	0.000	13:16	CALIBRATION CHECK RESULT 3 0.081		
Cal Check	0.081	13:16	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 2.5%		
Air Blank	0.000	13:17	SPREAD (MUST BE .005 OR LESS) 0.001		

**Pass**

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass		Air Blank	0.000	13:17
EEPROM Checksum Test	Pass		Subject Test	RFI*	13:17
Real Time Clock Test	Pass		Air Blank	0.000	13:18
DSP Test	Pass		*RFI Detect		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

**Pass**

**Pass**

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER			
SIGNATURE 		PRINT NAME STEVEN TRACY	
TYPE II PERMIT NUMBER 250175	EXPIRATION DATE 07/28/2017	TELEPHONE NUMBER 8168476250	



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**STEVEN K TRACY**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOXILYZER 5000, INTOXILYZER 8000**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/28/2015

NUMBER 250175

EXPIRES 7/28/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator TRACY, STEVEN  
 Permit No 250175  
 Date Issued 7/28/2015 Date Expires 7/28/2017