



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

**RECEIVED**

By Carol Day at 11:53 am, Feb 01, 2016

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN <b>931097</b>	NAME OF AGENCY <b>Lawson Police Department</b>	DATE OF INSPECTION <b>01-23-2016</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>1041 W Third Street Lawson MO 64062</b>		TIME OF INSPECTION <b>09:26</b>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <b>01/23/16 09:26</b>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <b>49</b> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER **Repro Marketing** LOT # **14001** EXP. DATE **04-30-16**

SIMULATOR TEMP (34°C ± 0.2°C) **34.0** °C SIMULATOR SN **DR5392** EXP. DATE **01-05-2017**

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 <b>0.098</b>	TEST 2 <b>0.098</b>	TEST 3 <b>0.098</b>
---------------------	---------------------	---------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <b>-</b>	(.0-.04) <b>-</b>	(.05-.09) <b>-</b>	(.10-.14) <b>2</b>	(.15-.19) <b>-</b>	OVER .19 <b>-</b>
-------------------	-------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

**Conforms with DoltSS standards**

**INSPECTING OFFICER**

SIGNATURE <b>Sgt Bruce Summa 201</b>	PRINT FULL NAME <b>Sgt Bruce Summa 201</b>
TYPE OF PERMIT NUMBER/EXPIRATION DATE <b>240362 10-03-2016</b>	TELEPHONE NUMBER <b>816-580-7210</b>

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901**

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**  
**LOT NUMBER: 14001**  
**EXPIRATION DATE: April 30, 2016 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 14001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1216 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 1, 2014  
The expiration date for this lot number is April 30, 2016 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
LAWSON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 901097  
01/23/16  
09:26

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM (04-07-2000): OKAY  
HEATERS  
SAMPLE CHAMBER: 49C  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST  
!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJKL  
MNOPQRSTUVWXYZ[\]^\_`abcdefshijklmno  
pqrstuvwxyz{|}~

Operator Signature *Sgt Bruce Summa 201*

2000

BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
LAWSON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 901097  
01/23/16

TESTING OFFICER:

SUMMA/BRUCE/A  
OFFICER I.D.: 201  
PERMIT NUMBER: 240362  
EXPIRATION DATE: 10/03/16  
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST .000 09:29  
INTERNAL STANDARD VERIFIED 09:29  
EXTERNAL STANDARD .099 09:30  
BLANK TEST .000 09:30  
EXTERNAL STANDARD .099 09:31  
BLANK TEST .000 09:32  
EXTERNAL STANDARD .099 09:32  
BLANK TEST .000 09:33

N = 3  
SIM. = .1  
AVG. = .099

Operator Signature *Sgt Bruce Summa 201*

2000

BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
LAWSON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 901097  
01/23/16

ARREST TIME: 09:00

SUBJECT NAME:  
RFI/TEST  
DOB: 11/23/45 SEX: M  
STATE/D.L.: MO/1234567  
ARRESTING OFFICER:  
SUMMA/BRUCE/A

OFFICER I.D.: 201  
TESTING OFFICER:  
SUMMA/BRUCE/A  
OFFICER I.D.: 201  
PERMIT NUMBER: 240362  
EXPIRATION DATE: 10/03/16  
MISCELLANEOUS DATA:  
RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST .000 09:36  
INTERNAL STANDARD VERIFIED 09:36  
RADIO INTERFERENCE

Operator Signature *Sgt Bruce Summa 201*

2000-02



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**BRUCE SUMMA**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/3/2014

NUMBER 240362

EXPIRES 10/3/2016

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator SUMMA, BRUCE  
 Permit No 240362  
 Date Issued 10/3/2014 Date Expires 10/3/2016