



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN <b>950135</b>	NAME OF AGENCY <b>Richland Police Dept.</b>	DATE OF INSPECTION <b>01/17/16</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>2015 Chestnut St. Richland, Mo 65556</b>		TIME OF INSPECTION <b>10:24</b>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <b>01/17/16 10:20</b>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <b>49</b> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER **QUINT LABORATORIES** LOT # **15220** EXP. DATE **9/28/17**

SIMULATOR TEMP (34°C ± 0.2°C) **34** °C SIMULATOR SN **DR6931** EXP. DATE **12-11-16**

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 = **0.102**      TEST 2 = **0.102**      TEST 3 = **0.102**

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	1	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

**This unit is operating within Dept. of Health rules & req. SIMULATOR CALIBRATED 12-11-2015**

**INSPECTING OFFICER**

SIGNATURE <b>David Moser</b>	PRINT FULL NAME <b>DAVID MOSER</b>
TYPE II PERMIT NUMBER/EXPIRATION DATE <b>250012 1/14/2017</b>	TELEPHONE NUMBER <b>573 765 4144</b>

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Blvd.  
Poplar Bluff, MO 63901**



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT  
TYPE II**

**DAVID L MOSER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/14/2015

NUMBER 250012

EXPIRES 1/14/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 590-0771 (9-10)

LAB-4 (95-10)

**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator MOSER, DAVID  
Permit No 250012  
Date Issued 1/14/2015 Date Expires 1/14/2017

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 15220 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on September 30, 2015, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1214% (w/vol) ethyl alcohol. The expiration date for this lot number is September 28, 2017 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

**NIST Traceability:**

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Face This Side Down - This Edge In First

**BAC DataMaster**  
Evidence Ticket

STATE OF MISSOURI  
RICHLAND POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 950195  
01/17/16


TESTING OFFICER:  
MOSEY/DAVID  
OFFICER I.D.: 406  
PERMIT NUMBER: 256012  
EXPIRATION DATE: 01/17/17  
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	10:26
INTERNAL STANDARD	VERIFIED	10:26
EXTERNAL STANDARD	.102	10:27
BLANK TEST	.000	10:27
EXTERNAL STANDARD	.102	10:28
BLANK TEST	.000	10:28
EXTERNAL STANDARD	.102	10:29
BLANK TEST	.000	10:29

N = 3  
SIM. = .1  
AVG. = .102

Operator Signature



Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
RICHLAND POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 950135  
01/17/16  
10:23

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

```
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFGHI
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrs
tuvwxyz{|}~*+,-./0123456789:;<=>?@ABCDEFGHI
```

Operator Signature *James W. [Signature]*

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
RICHLAND POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 950135  
01/17/16

ARREST TIME: 10:00  
SUBJECT NAME:  
JOE/N/DGE  
DOB: 09/16/54      SEX: M  
STATE/D.L.: MO/W576438765  
ARRESTING OFFICER:  
MOSE/DAVID  
OFFICER I.D.: 406  
TESTING OFFICER:  
SAME  
OFFICER I.D.: 406  
PERMIT NUMBER: 250012  
EXPIRATION DATE: 01/14/17  
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST      .000      10:53  
INTERNAL STANDARD      VERIFIED      10:53  
RADIO INTERFERENCE

Operator Signature

*David Moser*