



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 9:08 am, Jan 29, 2016

Complete this report at the time of the regular monthly preventive maintenance. Complete this report whenever the instrument is serviced or repaired and Retain the original and send a copy within 15 days to the Breath Alcohol Program.

DATAMASTER SN 950223	NAME OF AGENCY DIAMOND Police Department	DATE OF INSPECTION 01-28-2016
LOCATION OF INSTRUMENT (STREET AND CITY) 100 N. WASHINGTON Diamond, Mo. 64840		TIME OF INSPECTION 21:27

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 01-28-16 @ 21:29
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER **Guth Labs** LOT # **15120** EXP. DATE **04-29-17**

SIMULATOR TEMP (34°C ± 0.2°C) **34.0** °C SIMULATOR SN **SD1918** EXP. DATE **08-11-2009**

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .101	TEST 2 .102	TEST 3 .103
--------------------	--------------------	--------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <input checked="" type="checkbox"/> (0-.04) <input checked="" type="checkbox"/>	<input checked="" type="checkbox"/> (.05-.09) <input checked="" type="checkbox"/>	<input checked="" type="checkbox"/> (.10-.14) <input checked="" type="checkbox"/>	<input checked="" type="checkbox"/> (.15-.19) <input checked="" type="checkbox"/>	<input checked="" type="checkbox"/> OVER .19 <input checked="" type="checkbox"/>
--	---	---	---	--

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE Keith Brumfield	PRINT FULL NAME Keith Brumfield
TYPE II PERMIT NUMBER/EXPIRATION DATE 250046 02-20-2017	TELEPHONE NUMBER 417-499-1495

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-584-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 15120 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 4, 2015, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is April 29, 2017 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
KEITH J BRUMFIELD

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/20/2015

NUMBER 250046

EXPIRES 2/20/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BRUMFIELD, KEITH
 Permit No 250046
 Date Issued 2/20/2015 Date Expires 2/20/2017

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
DIAMOND POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 950220
01/29/16

TESTING OFFICER:
DRUMFIELD-KEITH/J
OFFICER I.D.: 87
PERMIT NUMBER: 250046
EXPIRATION DATE: 02/28/17
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	21:35
INTERNAL STANDARD	VERIFIED	21:35
EXTERNAL STANDARD	.101	21:35
BLANK TEST	.000	21:36
EXTERNAL STANDARD	.102	21:37
BLANK TEST	.000	21:37
EXTERNAL STANDARD	.103	21:38
BLANK TEST	.000	21:39

N = 3
SUM = .1
AVG = .102

Operator Signature

Keith Drumfield

2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
DIAMOND POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 950220
01/29/16
21:29

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~

Operator Signature

Keith Drumfield

2208-02

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
DIAMOND POLICE DEPARTMENT

DAC DATAMASTER SERIAL NUMBER 600220
01/20/10

ARREST TIME: 21:10
SUBJECT NAME:
R/F/I/TEST
DOB: 00/07/70 SEX: M
STATE/D.L.L.: MO/100450
ARRESTING OFFICER:
SAME
OFFICER I.D.: 07
TESTING OFFICER:
SAME
OFFICER I.D.: 07
PERMIT NUMBER: 258046
EXPIRATION DATE: 00/20/17
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLOW TEST	.000	21:40
INTERNAL STANDARD	VERIFIED	21:40
RADIO INTERFERENCE		

Operator Signature

[Handwritten Signature]