

WAS SENT TO US PRIOR TO
1/24/16
BUT RECEIVED LATE DUE TO
CLERICAL ISSUE ON OUR PART.
BML - 1/29/16

RECEIVED
By Brian Lutmer at 2:45 pm, Jan 29, 2016



MIS
STA
BR
SERVICES
DATAMASTER MAINTENANCE REPORT

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 950223	NAME OF AGENCY DIAMOND Police Department	DATE OF INSPECTION 12-29-2015
LOCATION OF INSTRUMENT (STREET AND CITY) 100 N. Washington Street Diamond, Mo. 64840		TIME OF INSPECTION 19:00

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 12-29-15 @ 19:00
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Guth Labs	LOT # 15120 EXP. DATE 04-29-17
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C	SIMULATOR SN SD1918 EXP. DATE 08-11-2009
<input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)	
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)	
<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE	
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE	
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE	
TEST 1 * .101	TEST 2 * .102
TEST 3 * .102	
<input checked="" type="checkbox"/> PERFORM R.F.I. TEST (PRINTOUT ATTACHED)	

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	<input checked="" type="checkbox"/> (0-.04)	<input checked="" type="checkbox"/> (.05-.09)	<input checked="" type="checkbox"/> (.10-.14)	<input checked="" type="checkbox"/> (.15-.19)	<input checked="" type="checkbox"/> OVER .19
----------	---	---	---	---	--

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS
(USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE Keith Brumfield	PRINT FULL NAME Keith Brumfield
TYPE II PERMIT NUMBER/EXPIRATION DATE 250046 02-20-2017	TELEPHONE NUMBER 417-499-1495

RETURN COMPLETED REPORT TO THE:
Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **15120** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 4, 2015**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1209%** (w/vol) ethyl alcohol. The expiration date for this lot number is **April 29, 2017** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

KEITH J BRUMFIELD

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/20/2015

NUMBER 250046

EXPIRES 2/20/2017

MO 680-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BRUMFIELD, KEITH
Permit No 250046
Date Issued 2/20/2015 Date Expires 2/20/2017

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
DIAMOND POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 950223
12/29/15

ARREST TIME: 19:00
SUBJECT NAME:
R/F/L/TEST
DOB: 09/07/72 SEX: M
STATE/D.L.: MO-12345678
ARRESTING OFFICER:
SAME
OFFICER I.D.: 87
TESTING OFFICER:
SAME
OFFICER I.D.: 97
PERMIT NUMBER: 250046
EXPIRATION DATE: 02/20/17
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	19:29
INTERNAL STANDARD	VERIFIED	19:29
RADIO INTERFERENCE		

Operator Signature

Keith Bumfield

220

Operator Signature

Keith Bumfield

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
DIAMOND POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 950223
12/29/15

TESTING OFFICER:
BRUMFIELD/KEITH/J
OFFICER I.D.: 87
PERMIT NUMBER: 250046
EXPIRATION DATE: 02/20/17
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	19:21
INTERNAL STANDARD	VERIFIED	19:21
EXTERNAL STANDARD	.101	19:21
BLANK TEST	.000	19:22
EXTERNAL STANDARD	.102	19:23
BLANK TEST	.000	19:23
EXTERNAL STANDARD	.102	19:24
BLANK TEST	.000	19:25

N = 3
SIM. = .1
AVG. = .1016

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
DIAMOND POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 950829
12/29/15
19:00

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49:
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#%&'()*+,-./:0123456789:;@#%&'()*+,-./:0123456789:;
ABCDEFGHIJKLMNPOQRSTUVWXYZ[]^_`abcedefghijklmnop
qrstuvwxyz{|}~*

Operator Signature

