

Jan. 13. 2016



MISSOURI  
STATE  
BREATH  
DATAMASTER

**WAS SENT TO US PRIOR TO  
1/24/16  
BUT RECEIVED LATE DUE TO  
CLERICAL ISSUE ON OUR PART.  
BML - 1/29/16**

SERVICES

**RECEIVED**

No. 0226 P. 2

By Brian Lutmer at 2:45 pm, Jan 29, 2016

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 960062	NAME OF AGENCY Riverview PD	DATE OF INSPECTION 01/13/2016
LOCATION OF INSTRUMENT (STREET AND CITY) 9699 Lilac Riverview		TIME OF INSPECTION 9:26 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>01/13/16 09:26</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>49</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Guth LOT # 15120 EXP. DATE 04/29/2017

SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN SD2769 EXP. DATE 06/26/2016

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 <u>.099</u>	TEST 2 <u>.100</u>	TEST 3 <u>.100</u>
--------------------	--------------------	--------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

**INSPECTING OFFICER**

SIGNATURE 	PRINT FULL NAME Nicholes C. Allred
TYPE II PERMIT NUMBER/EXPIRATION DATE 250001 01/02/2017	TELEPHONE NUMBER (314) 868-9130

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Blvd.  
Poplar Bluff, MO 63901

**BAC DataMaster**  
**Evidence Ticket**

STATE OF MISSOURI  
RIVERVIEW POLICE DEPARTMENT


BAC DATAMASTER SERIAL NUMBER 960362  
01/13/16

ARREST TIME: 09:05  
SUBJECT NAME:  
KROENKE/ENDS/STANLEY  
DOB: 02/11/22 SEX: M  
STATE/D.L.L.# NO/056503656  
ARRESTING OFFICER:  
ALLRED/NICHOLS/C  
OFFICER I.D.# 247  
TESTING OFFICER:  
ALLRED/NICHOLS C  
OFFICER I.D.# 247  
PERMIT NUMBER: 257001  
EXPIRATION DATE: 01/02/17  
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	09:29
INTERNAL STANDARD	VERIFIED	09:39
SUBJECT SAMPLE	.000	09:59
RADIO INTERFERENCE		

Operator Signature



**GUTH LABORATORIES, INC.**

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 15120 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 4, 2015, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is April 29, 2017 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

**BAC DataMaster**  
**Evidence Ticket**

STATE OF MISSOURI  
RIVERVIEW POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 958862  
01/13/16

TESTING OFFICER:  
ALLRED/NICHOLS/C  
OFFICER I.D.: 247  
PERMIT NUMBER: 250001  
EXPIRATION DATE: 01/22/17  
MISCELLANEOUS DATA:

----- SUPERVISOR MODE -----

BLANK TEST	.000	00:00
INTERNAL STANDARD	VERIFIED	00:00
EXTERNAL STANDARD	.000	00:00
BLANK TEST	.000	00:01
EXTERNAL STANDARD	.100	00:02
BLANK TEST	.000	00:03
EXTERNAL STANDARD	.100	00:03
BLANK TEST	.000	00:04

N = 3  
SIM. = .1  
AVG. = .0006

Operator Signature



**BAC DataMaster**  
Evidence Ticket

STATE OF MISSOURI  
RIVERVIEW POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 980062  
01/13/16  
09:26

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;@=?`~{|}~  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefg|hijklmno  
pqrstuvwxyz{|}~`~

Operator Signature





STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**

**NICHOLAS C ALLRED**

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/2/2015

NUMBER 250001

EXPIRES 1/2/2017

MO 590-0771 (8-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator ALLRED, NICHOLAS  
Permit No 250001  
Date Issued 1/2/2016 Date Expires 1/2/2017