



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 035484	NAME OF AGENCY Saint Charles County Police Department	DATE OF INSPECTION 01/14/2023
LOCATION OF INSTRUMENT (STREET AND CITY) 301 North Second Street, Saint Charles, MO 63301		TIME OF INSPECTION 8:41 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG109701 EXP. DATE 04/07/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1  .082

TEST 2  .081

TEST 3  .081

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE

PRINT NAME  
 Mike Johnson

TYPE II PERMIT NUMBER/EXPIRATION DATE  
 220150 / 05-25-2024

TELEPHONE NUMBER  
 (636) 949-3000

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 800484  
Version no: 5320

TEST RECORD 80263

Temp Date Time <sup>g/</sup> 210L

Air Blank  
01/14/73 08:44 .000  
Calibration Check:  
19 01/14/73 08:44 .002

Subject Name

Test #1

Subject I.D.

Operator Name: I.D.

Cpl. Johnson 220150

Location

SecDoc

AS IV Serial no: 800484  
Version no: 5320

TEST RECORD 80264

Temp Date Time <sup>g/</sup> 210L

Air Blank  
01/14/73 08:44 .000  
Calibration Check:  
20 01/14/73 08:44 .001

Subject Name

Test #2

Subject I.D.

Operator Name: I.D.

Cpl. Johnson 220150

Location

SecDoc

AS IV Serial no: 800484  
Version no: 5320

TEST RECORD 80265

Temp Date Time <sup>g/</sup> 210L

Air Blank  
01/14/73 08:45 .000  
Calibration Check:  
21 01/14/73 08:45 .001

Subject Name

Test #3

Subject I.D.

Operator Name: I.D.

Cpl. Johnson 220150

Location

SecDoc

TEST RECORD 80266 <sup>g/</sup>  
Temp Date Time 210L  
NOTE NO  
12 01/14/73 08:44  
Subject Name  
LTI Test  
Subject I.D.  
Operator Name: I.D.  
Cpl. Johnson 220150  
Location  
SecDoc

TEST RECORD 80267 <sup>g/</sup>  
Temp Date Time 210L  
Air Blank  
01/14/73 08:47 .000  
Subject Name: Auto  
22 01/14/73 08:47 .000  
Subject Name  
SME-Test  
Subject I.D.  
Operator Name: I.D.  
Cpl. Johnson 220150  
Location  
SecDoc

AS IV Serial no: 800484  
Version no: 5320

AS IV Serial no: 800484  
Version no: 5320



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
*Exclusive Supplier*  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 7-Apr-2021

**Lot # AG109701 Model 108cacc**

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
7-Apr-2023	108	Ethanol Nitrogen	0.082 ± 0.002 BrAC (223 ppm) Balance

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

  

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

**Analytical Method:**      **NDIR**

Digitally signed by Quality Control  
 Date: 2021.04.07 17:42:01 -05:00  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)

Approved for Release: \_\_\_\_\_

*Rod Marsala*  
 \_\_\_\_\_  
 Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**MIKE JOHNSON**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/25/2022

NUMBER 220150

EXPIRES 5/25/2024

*Mike Mason*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*David J. Nielson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** JOHNSON, MIKE  
**Permit No** 220150  
**Date Issued** 5/25/2022 **Date Expires** 5/25/2024

