

AS IV Serial no: 035714
Version no: 532B

TEST RECORD 01509
Temp Date Time 210L s/

Air Blank:
11/21/23 19:02 .000
Calibration Check:
22 11/21/23 19:02 .101

Subject Name
Blank

Subject I.D.
114

Operator Name, I.D.
Lackey 114

Location
203 Johnsons Memorial Park

Wright City MO 63390

AS IV Serial no: 035714
Version no: 532B

TEST RECORD 01510
Temp Date Time 210L s/

Air Blank:
11/21/23 19:05 .000
Calibration Check:
23 11/21/23 19:05 .102

Subject Name
Blank

Subject I.D.
114

Operator Name, I.D.
Lackey 114

Location
203 VMP

Wright City

AS IV Serial no: 035714
Version no: 532B

TEST RECORD 01511
Temp Date Time 210L s/

Air Blank:
11/21/23 19:07 .000
Calibration Check:
24 11/21/23 19:07 .102

Subject Name
Blank

Subject I.D.
114

Operator Name, I.D.
Lackey 114

Location
203 VMP

Wright City

AS IV Serial no: 035714
Version no: 532B

TEST RECORD 01512
Temp Date Time 210L s/

VOID: REI
12 11/21/23 19:09

Subject Name
Blank

Subject I.D.
114

Operator Name, I.D.
Lackey 114

Location
203 VMP

Wright City



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 23390 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2023, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1207% (w/vol) ethyl alcohol. The expiration date for this lot number is October 17, 2025 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

MATT LACKEY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/23/2022

Laura A. Way

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 220101

David J. Nickelson

EXPIRES 3/23/2024

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 **STATE OF MISSOURI**
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator LACKEY, MATT
Permit No 220101
Date Issued 3/23/2022 Date Expires 3/23/2024

