



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 9:44 pm, Dec 30, 2023

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 042311	PRINTER SN 099.3586.572	DATE OF INSPECTION 12/28/2023
LOCATION OF INSTRUMENT (STREET AND CITY) SCCPD / 101 Sheriff Dierker Court, O'Fallon, Missouri 63366		TIME OF INSPECTION 9:28 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Intoximeters LOT # AG313002 EXP. DATE 05/10/2025
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .103	TEST 2 .103	TEST 3 .102
--------------	--------------	--------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME Piant Jr, Jeffery
TYPE # PERMIT NUMBER/EXPIRATION DATE 230051/03/27/2025	TELEPHONE NUMBER (636) 949-3000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

AS IV Serial no: 042311
Version no: 532C

TEST RECORD 00557

Temp Date Time ^{g/} 210L

Air Blank:
12/28/23 21:28 .000
Calibration Check:
21 12/28/23 21:28 .103

Subject Name

Test

Subject I.D.

1

Operator Name, I.D.

Piant 230051

Location

SCCPD

AS IV Serial no: 042311
Version no: 532C

TEST RECORD 00558

Temp Date Time ^{g/} 210L

Air Blank:
12/28/23 21:30 .000
Calibration Check:
21 12/28/23 21:30 .103

Subject Name

Test

Subject I.D.

2

Operator Name, I.D.

Piant 230051

Location

SCCPD

AS IV Serial no: 042311
Version no: 532C

TEST RECORD 00559

Temp Date Time ^{g/} 210L

Air Blank:
12/28/23 21:32 .000
Calibration Check:
22 12/28/23 21:32 .102

Subject Name

Test 3

Subject I.D.

3

Operator Name, I.D.

Piant 230051

Location

SCCPD

AS IV Serial no: 042311
Version no: 532C

TEST RECORD 00560

Temp Date Time ^{g/} 210L

VOID: RFI
12 12/28/23 21:34

Subject Name

Test

Subject I.D.

RFI

Operator Name, I.D.

Piant 230051

Location

SCCPD



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 11-May-2023

Lot # AG313002 Model 108

Exp Date	Cyl. Type	Component	Certified Concentration
10-May-2025	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm)

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by: Quality Control
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)
 Date: 05.11.2023 21:07

Approved for Release: _____

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

JEFFERY PIANT JR.

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field services and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

This determination of the appropriateness of this permit is based on the information furnished under the provisions of sections 57.120 through 57.124, RSMo and 506.171 through 506.174 RSMo.

DATE: 3/27/2023

NUMBER 230051

EXPIRES 3/27/2025

M. M. M...

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David S. Nielsen

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB 10/10/07

MO-500 (07/16-00)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator PIANT JR., JEFFERY

Permit No 230051

Date Issued 3/27/2023 . Date Expires 3/27/2025.

