

AS IV Serial no: 087965
Version no: 532B

TEST RECORD 00636

Temp Date Time ^{s/} 210L

Air Blank:
03/01/23 14:08 .000
Calibration Check:
19 03/01/23 14:08 .100

Subject Name

Test 1

Subject I.D.

Operator Name, I.D.

Rew503 220231 9/9/24

Location

2501 W. Mechanic

AS IV Serial no: 087965
Version no: 532B

TEST RECORD 00637

Temp Date Time ^{s/} 210L

Air Blank:
03/01/23 14:10 .000
Calibration Check:
20 03/01/23 14:10 .099

Subject Name

Test 2

Subject I.D.

Operator Name, I.D.

Rew503 220231 9/9/24

Location

2501 W. Mechanic

AS IV Serial no: 087965
Version no: 532B

TEST RECORD 00638

Temp Date Time ^{s/} 210L

Air Blank:
03/01/23 14:12 .000
Calibration Check:
20 03/01/23 14:12 .100

Subject Name

Test 3

Subject I.D.

Operator Name, I.D.

Rew503 220231 9/9/24

Location

2501 W. Mechanic

AS IV Serial no: 087965
Version no: 532B

TEST RECORD 00639

Temp Date Time ^{s/} 210L

VOID: RFI
12 03/01/23 14:14

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

Rew 503 220231 9/9/24

Location

2501 W. Mechanic

AS IV Serial no: 087965
Version no: 532B

TEST RECORD 00640

Temp Date Time ^{s/} 210L

Air Blank:
03/01/23 14:15 .000
Subject Test: Auto
21 03/01/23 14:15 .000

Subject Name

Test Blank

Subject I.D.

Operator Name, I.D.

Rew503 220231 9/9/24

Location

2501 W. Mechanic



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



**PERMIT
 TYPE II**

JAMES N. REW

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/9/2022

NUMBER 220231

EXPIRES 9/9/2024

Laura E. Way

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Donald A. Kamm

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Cralg Road
 St. Louis, Mo 63146

Test Date: 14-Dec-2021

Lot # AG134807 Model 34

Exp Date 14-Aug-2023	Cyl. Type 34	Component Ethanol Nitrogen	Certified Concentration 0.100 ± 2% BrAC (272 ppm)
--------------------------------	------------------------	---	---

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by:Quality Control
 Reason:Dry gas standard certification of analysis
 Location:Airgas USA LLC (Lab)
 Date:12.18.2021 14:38

Approved for Release: *Rod Marsala*
 Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07