

CARROLLTON POLICE DEPARTMENT-BREATH ALCOHOL PROGRAM

AS-IV Serial no: 087971
Version no: 532B

TEST RECORD 00424
Temp Date Time 210L

Air Blank: 08/01/23 19:55 .000
Calibration Checks: 24 08/01/23 19:55 .000

Subject Name
TEST #1
Subject I.D.

Operator Name, I.D. #
WILLIAM BARGER #230029
Location
1320 U.S. 65 N

CARROLLTON MO
64633

TEST# 1

AS-IV Serial no: 087971
Version no: 532B

TEST RECORD 00427
Temp Date Time 210L

Air Blank: 08/01/23 20:00 .000
Calibration Checks: 24 08/01/23 20:00 .000

Subject Name
TEST # 2
Subject I.D.

Operator Name, I.D. #
WILLIAM BARGER #230029
Location
1320 U.S. 65 N

CARROLLTON, MO
64633

TEST #2

AS-IV Serial no: 087971
Version no: 532B

TEST RECORD 00428
Temp Date Time 210L

Air Blank: 08/01/23 20:05 .000
Calibration Checks: 25 08/01/23 20:05 .000

Subject Name
TEST #3
Subject I.D.

Operator Name, I.D. #
WILLIAM BARGER #230029
Location
1320 U.S. 65 N

CARROLLTON, MO
64633

TEST #3

AS-IV Serial no: 087971
Version no: 532B

TEST RECORD 00429
Temp Date Time 210L

Air Blank: 08/01/23 20:07

Subject Name
REF TEST
Subject I.D.

Operator Name, I.D. #
WILLIAM BARGER #230029
Location
1320 U.S. 65 N

CARROLLTON, MO
64633

RFI TEST

AS-IV S/N: 087971 Type II Permit Holder: William Barger #230029-Exp: 02/17/2025 Date of Maintenance Report: 8/1/2023



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Test Date: 9-Feb-2023

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Lot # AG304002 Model 108

Exp Date	Cyl. Type	Component	Certified Concentration
9-Feb-2025	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm)


Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by:Quality Control
 Reason:Dry gas standard certification of analysis
 Location:Airgas USA LLC (Lab)
 Date:02.09.2023 19:01

Approved for Release: 
 Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
WILLIAM BARGER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/17/2023

NUMBER 230029

EXPIRES 2/17/2025

MO 580-0771 (6-10)

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula F. Nielson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BARGER, WILLIAM
 Permit No 230029
 Date Issued 2/17/2023 Date Expires 2/17/2025

