



AS IV Serial no: 087972  
Version no: 532B

TEST RECORD 00709

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
11/04/23 22:10 .000  
Calibration Check:  
19 11/04/23 22:10 .105

Subject Name

Test #1

Subject I.D.

NA

Operator Name, I.D.

Sgt D. Estrada #3072

Location

232 Vance Rd  
(7th PCT)

AS IV Serial no: 087972  
Version no: 532B

TEST RECORD 00710

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
11/04/23 22:13 .000  
Calibration Check:  
20 11/04/23 22:13 .104

Subject Name

Test #2

Subject I.D.

NA

Operator Name, I.D.

Sgt D. Estrada #3070

Location

232 Vance Rd  
(7th PCT)

AS IV Serial no: 087972  
Version no: 532B

TEST RECORD 00711

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
11/04/23 22:15 .000  
Calibration Check:  
21 11/04/23 22:15 .103

Subject Name

Test #3

Subject I.D.

NA

Operator Name, I.D.

Sgt D. Estrada #3072

Location

232 Vance Rd  
(7th PCT)

AS IV Serial no: 087972  
Version no: 532B

TEST RECORD 00712

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 11/04/23 22:19

Subject Name

RFI

Subject I.D.

NA

Operator Name, I.D.

Sgt D. Estrada #3072

Location

232 Vance Rd  
(7th PCT)



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



PERMIT  
TYPE II

DAWN M. ESTRADA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

EXPIRES 4/2/2023

PERMIT NUMBER 230059

ISSUED 4/2/2025

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES