



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 087974	NAME OF AGENCY Oregon County sheriff's Office	DATE OF INSPECTION 11/09/2023
LOCATION OF INSTRUMENT (STREET AND CITY) Oregon County Sheriff's Office		TIME OF INSPECTION 11:57 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Lab LOT # 23180 EXP. DATE 05/17/2025

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIM. SN Mp6537 SIM. NIST EXP DATE 05/23/2024

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)


- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .099	TEST 2 ← .098	TEST 3 ← .098
---------------	---------------	---------------


RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use of  if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME J.A. Martin DSN 4064
TYPE II PERMIT NUMBER/EXPIRATION DATE 230107 05/30/2025	TELEPHONE NUMBER (417) 778-6611

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 087974
Version no: 532B

TEST RECORD 00280

Temp	Date	Time	g/ 210L
------	------	------	------------

Air Blank:
11/09/23 11:57 .000
Calibration Check:
23 11/09/23 11:57 .099

Subject Name

Test 1

Subject I.D.

Maintenance

Operator Name, I.D.

J. A. M. 4064

Location

Oregon Co. Sheriff's Dept

AS IV Serial no: 087974
Version no: 532B

TEST RECORD 00281

Temp	Date	Time	g/ 210L
------	------	------	------------

Air Blank:
11/09/23 11:59 .000
Calibration Check:
24 11/09/23 11:59 .098

Subject Name

Test 2

Subject I.D.

Maintenance

Operator Name, I.D.

J. A. M. 4064

Location

Oregon Co. Sheriff's Dept

AS IV Serial no: 087974
Version no: 532B

TEST RECORD 00282

Temp	Date	Time	g/ 210L
------	------	------	------------

Air Blank:
11/09/23 12:01 .000
Calibration Check:
25 11/09/23 12:01 .098

Subject Name

Test 3

Subject I.D.

Maintenance

Operator Name, I.D.

J. A. M. 4064

Location

Oregon Co. Sheriff's Dept

AS IV Serial no: 087974
Version no: 532B

TEST RECORD 00283

Temp	Date	Time	g/ 210L
------	------	------	------------

VOID: RFI
12 11/09/23 12:03

Subject Name

R. F. I

Subject I.D.

Maintenance

Operator Name, I.D.

J. A. M. 4064

Location

Oregon Co. Sheriff's Dept



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 23180 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 18, 2023, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1220% (w/vol) ethyl alcohol. The expiration date for this lot number is May 17, 2025 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN11172002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

JAMES A. MARTIN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/30/2023

NUMBER 230107

EXPIRES 5/30/2025

Mike Massum

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David F. Nickelson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MARTIN, JAMES
Permit No 230107
Date Issued 5/30/2023 Date Expires 5/30/2025

