



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 095961	PRINTER SN 099.3586.643	DATE OF INSPECTION 01/22/2023
LOCATION OF INSTRUMENT (STREET AND CITY) 715 W Center Circle, Nixa MO		TIME OF INSPECTION 0:01 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER ACS LOT # 202205A EXP. DATE 05/17/2024

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN MP5537 SIMULATOR EXP DATE 07/11/2023

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➡ .098	TEST 2 ➡ .100	TEST 3 ➡ .097
---------------	---------------	---------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
Meets and exceeds standards.

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Sgt R Seiner - Nixa PD
TYPE II PERMIT NUMBER/EXPIRATION DATE 220217 09/02/2024	TELEPHONE NUMBER (417) 725-2510

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

Nixa Police Department

Calibration Check slip's

AS IV Serial no: 095961
Version no: 532B

TEST RECORD 02603

Temp Date Time 210L

Air Blank:
01/22/23 06:28 .000

Calibration Check:
20 01/22/23 06:28 .098

Subject Name
TEST

Subject I.D.

Operator Name, I.D.
RSEINER 220217

Location
NIXA PD

AS IV Serial no: 095961
Version no: 532B

TEST RECORD 02607

Temp Date Time 210L

Air Blank:
01/22/23 06:37 .000

Calibration Check:
24 01/22/23 06:37 .097

Subject Name
TEST

Subject I.D.

Operator Name, I.D.
RSEINER 220217

Location
NIXA PD

AS IV Serial no: 095961
Version no: 532B

TEST RECORD 02604

Temp Date Time 210L

Air Blank:
01/22/23 06:30 .000

Calibration Check:
22 01/22/23 06:30 .100

Subject Name
TEST

Subject I.D.

Operator Name, I.D.
RSEINER 220217

Location
NIXA PD

Nixa Police Department

Blank (Zero) test Evidence slip

AS IV Serial no: 095961			
Version no: 532B			
TEST RECORD 02606			
Temp	Date	Time	9/ 210L

Air Blank:			
	01/22/23	06:35	.000
Subject Test: Auto			
24	01/22/23	06:35	.000

Subject Name			
TEST TEST			

Subject I.D.			

Operator Name, I.D.			
DSENER 270217			

Location			
NIXA PD NIXA PD			

Nixa Police Department

RFI Evidence slip

AS IV Serial no: 095961			
Version no: 532B			
TEST RECORD 02608			
Temp	Date	Time	s/ 210L

VOID: RFI			
12 01/22/23 06:38			

Subject Name			
TEST			

Subject I.D.			

Operator Name, I.D.			
REINER 220217			

Location			
NIXA PD			

Reference Material Certificate of Analysis

Product identification:	Alcohol Reference Solution (ARS)
CRM:	Ethanol in aqueous solution
Reference concentration:	<u>1.2100</u> g/L
Analytical concentration:	<u>1.2153</u> g/L
Batch size:	<u>2291</u> bottles
Lot number:	<u>202205A</u>
Date of production:	<u>2022.05.17</u> (yyyy.mm.dd)


This certificate covers the production, analysis and certification of the Certified Reference Material (CRM) manufactured by Alcohol Countermeasure Systems Corp (ACS), located at 60 International Boulevard, Toronto, ON M9W 6J2 CANADA.

This metrology laboratory is accredited in accordance with international standards ISO/IEC 17025:2017 for competence of testing and calibration laboratories, and ISO/IEC 17034:2016 for the competence of reference material producers, demonstrating technical competence within the defined scope and the operation of a laboratory quality management system.

The analysis for certification of each lot of ARS is conducted within ACS forensic laboratory by an independent chemist using the direct injection, flame ionization gas chromatographic procedure coupled with the internal standard technique commensurate with forensic alcohol analysis.

This certificate is valid only for the ARS lot number identified above and does not extend to any other production lot of alcohol reference solution.

This certificate may only be reproduced in full.



For and on behalf of the company,
ALCOHOL COUNTERMEASURE
SYSTEMS CORP.

Felix JE Comeau, B.Sc (Honours)
Laboratory Director



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



**PERMIT
 TYPE II**

ROBERT A. SEINER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/2/2022

NUMBER 220217

EXPIRES 9/2/2024

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David J. Nickelson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator SEINER, ROBERT
 Permit No 220217
 Date Issued 9/2/2022 Date Expires 9/2/2024

