

TEST REPORT 00451

Temp Date Time 2

Air Blower 08/12/23 10:09 .00

Calibration Check: 19 08/12/23 10:09 .00

Subject Name

MAINT. CHECK

Subject I.D.

Operator Name: I.D.

S.K. TALBURT #1110

Location

TROOP E HQ

POPLAR BLUFF, MO

AS IU Serial not 892400
Version not 532B

TEST REPORT 00453

Temp Date Time 2100

Air Blower 08/12/23 10:15 .000

Calibration Check: 21 08/12/23 10:15 .100

Subject Name

MAINT CHECK

Subject I.D.

Operator Name: I.D.

S.K. TALBURT #1110

Location

TROOP E HQ

POPLAR BLUFF, MO

AS IU Serial not 892400
Version not 532B

TEST REPORT 00452

Temp Date Time 2100

Air Blower 08/12/23 10:12 .00

Calibration Check: 20 08/12/23 10:12 .100

Subject Name

MAINT CHECK

Subject I.D.

Operator Name: I.D.

S.K. TALBURT #1110

Location

TROOP E HQ

POPLAR BLUFF, MO

AS IU Serial not 892400
Version not 532B

TEST REPORT 00454

Temp Date Time 2100

Air Blower 12 08/12/23 10:15

Subject Name

MAINT CHECK

Subject I.D.

Operator Name: I.D.

S.K. TALBURT #1110

Location

TROOP E HQ

POPLAR BLUFF, MO



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23180** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1220%** (w/vol) ethyl alcohol. The expiration date for this lot number is **May 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN1172002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
SHAYNE K. TALBURT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/7/2021

NUMBER 210293

EXPIRES 12/7/2023

Laura E. Way

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Donald A. Kamm

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator TALBURT, SHAYNE
Permit No 210293
Date Issued 12/7/2021 Date Expires 12/7/2023

