



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 BY: Tracy Crowe at 2:42 pm, 6/28/23

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097403/099.3586.830	NAME OF AGENCY Shrewsbury PD	DATE OF INSPECTION 06/28/2023
LOCATION OF INSTRUMENT (STREET AND CITY) 4400 Shrewsbury Ave, Shrewsbury		TIME OF INSPECTION 8:00 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION
- COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters Inc LOT # AG211501 EXP. DATE 04/25/2025

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .083 TEST 2 .082 TEST 3 .082

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument functioning according to DHSS rules and regulations.

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Zachery King
TYPE II PERMIT NUMBER/EXPIRATION DATE 210209-09/14/2023	TELEPHONE NUMBER (314) 647-5656

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

2

AS IV Serial no: 097403
Version no: 532B

TEST RECORD 00854
Temp Date Time 210L
s/

Air Blank: 06/28/23 20:05 .000
Calibration Check: 22 06/28/23 20:05 .082

Subject Name
TESTER
Subject I.D. 123

Operator Name, I.D.
PO KING 123
Location SPD

AS IV Serial no: 097403
Version no: 532B

TEST RECORD 00853
Temp Date Time 210L
s/

Air Blank: 06/28/23 20:03 .000
Calibration Check: 22 06/28/23 20:03 .083

Subject Name
TESTER
Subject I.D. 123

Operator Name, I.D.
PO KING 123
Location SPD

3

AS IV Serial no: 097403
Version no: 532B

TEST RECORD 00855
Temp Date Time 210L
s/

Air Blank: 06/28/23 20:07 .000
Calibration Check: 23 06/28/23 20:07 .082

Subject Name
TESTER
Subject I.D. 123

Operator Name, I.D.
PO KING 123
Location SPD

RFI!

AS IV Serial no: 097403
Version no: 532B

TEST RECORD 00856
Temp Date Time 210L
s/

VOID: RFI
12 06/28/23 20:10

Subject Name
TESTER
Subject I.D. 123

Operator Name, I.D.
PO KING 123
Location SPD



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Test Date: 25-Apr-2022

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Lot # AG211501 Model 108

Exp Date	Cyl. Type	Component	Certified Concentration
25-Apr-2024	108	Ethanol Nitrogen	0.080 ± 0.002 BrAC (218 ppm)

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by:Quality Control
 Reason:Dry gas standard certification of analysis
 Location:Airgas USA LLC (Lab)
 Date:04.29.2022 15:19

Approved for Release: Rod Marsala
 Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
ZACHERY KING

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/14/2021

Laura A. Day
 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 210209

EXPIRES 9/14/2023

Donald A. Kamm
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator KING, ZACHERY
 Permit No 210209
 Date Issued 9/14/2021 Date Expires 9/14/2023

