

1

AS IV Serial no: 097403
Version no: 532B

TEST RECORD 00942

Temp Date Time 210L
9/

Air Blank:
10/05/23 18:20 .000

Calibration Check:
20 10/05/23 18:20 .078

Subject Name
TESTER

Subject I.D.
123

Operator Name, I.D.
PO KING 125

Location
SPD

2

AS IV Serial no: 097403
Version no: 532B

TEST RECORD 00943

Temp Date Time 210L
9/

Air Blank:
10/05/23 18:22 .000

Calibration Check:
21 10/05/23 18:22 .078

Subject Name
TESTER

Subject I.D.
123

Operator Name, I.D.
PO KING 125

Location
SPD

3

AS IV Serial no: 097403
Version no: 532B

TEST RECORD 00944

Temp Date Time 210L
9/

Air Blank:
10/05/23 18:24 .000

Calibration Check:
22 10/05/23 18:24 .077

Subject Name
TESTER

Subject I.D.
123

Operator Name, I.D.
PO KING 125

Location
SPD

RFI

AS IV Serial no: 097403
Version no: 532B

TEST RECORD 00945

Temp Date Time 210L
9/

VOID: RFI
12 10/05/23 18:25

Subject Name
TESTER

Subject I.D.
123

Operator Name, I.D.
PO KING 125

Location
SPD



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 25-Apr-2022

Lot # AG211501 **Model** 108

Exp Date 25-Apr-2024	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0.080 ± 0.002 BrAC (218 ppm)
--------------------------------	-------------------------	---	--

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by:Quality Control
 Reason: Dry gas standard certification of analysis
 Location:Airgas USA LLC (Lab)
 Date:04.28.2022 15:19

Approved for Release: _____

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
ZACHERY KING

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/7/2023

NUMBER 230200

EXPIRES 9/7/2025

Mike Massum

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nickelson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator KING, ZACHERY
Permit No 230200
Date Issued 9/7/2023 **Date Expires** 9/7/2025

