



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

By Tracy Crews at 9:12 am, Aug 25, 2023

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097416	NAME OF AGENCY Greene County Sheriff's Office	DATE OF INSPECTION 08/24/2023
LOCATION OF INSTRUMENT (STREET AND CITY) 5100 W. Division Springfield, MO 65802		TIME OF INSPECTION 12:29 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Intoximeters LOT # AG306807 EXP. DATE 03/09/2025
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE
- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .095	TEST 2 .096	TEST 3 .097
--------------------	--------------------	--------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE <i>Ronald L. Killingsworth</i>	PRINT NAME Ronald L. Killingsworth
TYPE II PERMIT NUMBER/EXPIRATION DATE 220287 / 12-23-24	TELEPHONE NUMBER (417) 829-6216

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 097416
Version no: 532B

TEST RECORD 01001

Temp	Date	Time	s/ 210L
Air Blank:			
	08/24/23	12:29	.000
Calibration Check:			
	28 08/24/23	12:29	.095

Subject Name

Killingsworth

Subject I.D.

220287/12-23-24

Operator Name, I.D.

Test #1

Location

GC50

AS IV Serial no: 097416
Version no: 532B

TEST RECORD 01002

Temp	Date	Time	s/ 210L
Air Blank:			
	08/24/23	12:31	.000
Calibration Check:			
	28 08/24/23	12:31	.096

Subject Name

Killingsworth

Subject I.D.

220287/12-23-24

Operator Name, I.D.

Test #2

Location

GC50

AS IV Serial no: 097416
Version no: 532B

TEST RECORD 01003

Temp	Date	Time	s/ 210L
Air Blank:			
	08/24/23	12:33	.000
Calibration Check:			
	29 08/24/23	12:33	.097

Subject Name

Killingsworth

Subject I.D.

220287/12-23-24

Operator Name, I.D.

Test #3

Location

GC50

AS IV Serial no: 097416
Version no: 532B

TEST RECORD 01004

Temp	Date	Time	s/ 210L
VOID: RFI			
	12 08/24/23	12:34	

Subject Name

Killingsworth

Subject I.D.

220287/12-23-24

Operator Name, I.D.

RFI Test

Location

GC50

AS IV Serial no: 097416
Version no: 532B

TEST RECORD 01005

Temp	Date	Time	s/ 210L
Air Blank:			
	08/24/23	12:36	.000
Subject Test: Auto			
	29 08/24/23	12:36	.000

Subject Name

Killingsworth

Subject I.D.

220287/12-23-24

Operator Name, I.D.

Blank

Location

GC50



Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 13-Mar-2023

Lot # AG306807 Model 108

Exp Date 9-Mar-2025	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0.100 ± 2% BrAC (272 ppm)
------------------------	------------------	----------------------------------	--

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by:Quality Control
Reason:Dry gas standard certification of analysis
Location:Airgas USA LLC (Lab)
Date:03.16.2023 13:02

Approved for Release:

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

RONALD L. KILLINGSWORTH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/23/2022

NUMBER 220287

EXPIRES 12/23/2024

Mike Massum

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David J. Nielson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (RS-10)

 STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator KILLINGSWORTH, RONALD
Permit No 220287
Date Issued 12/23/2022 Date Expires 12/23/2024

