



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

By Tracy Crews at 9:21 am, Mar 13, 2023

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097438	NAME OF AGENCY Kansas City Police Department	DATE OF INSPECTION 03-05-23
------------------------------------	--	---------------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 9701 MARION PARK DR, KANSAS CITY	TIME OF INSPECTION 0330
---	-----------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER INTOXIMETERS LOT # AG132803 EXP. DATE 11/24/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIM. SN _____ SIM. NIST EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1	.080	TEST 2	.080	TEST 3	.080
--------	-------------	--------	-------------	--------	-------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09) 1	(.10-.14)	(.15-.19)	(OVER .19) 2
----------	---------	--------------------	-----------	-----------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument meets all DOHSS standards and guidelines.

INSPECTING OFFICER

SIGNATURE P.O. J. Infranca

PRINT NAME **P.O. Infranca 5670**

TYPE II PERMIT NUMBER/EXPIRATION DATE **210128 06/21/2023**

TELEPHONE NUMBER **() 816-382-5897**

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

Completed 3.05.2023
0330 hours.

AS IV Serial no: 097438
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00370

Temp Date Time ^{a/} 210L

Air Blank:
03/05/23 03:24 .000
Calibration:
27 03/05/23 03:24 .000

Subject Name

Test #1

Subject I.D.

Operator Name, I.D.

Infranca 210128

Location

KCPD Traffic

AS IV Serial no: 097438
Version no: 532B

TEST RECORD 00371

Temp Date Time ^{a/} 210L

Air Blank:
03/05/23 03:27 .000
Calibration Check:
27 03/05/23 03:27 .000

Subject Name

Test #2

Subject I.D.

Operator Name, I.D.

Infranca 210128

Location

KCPD Traffic

AS IV Serial no: 097438
Version no: 532B

TEST RECORD 00372

Temp Date Time ^{a/} 210L

Air Blank:
03/05/23 03:29 .000
Calibration Check:
28 03/05/23 03:29 .000

Subject Name

Test #3

Subject I.D.

Operator Name, I.D.

Infranca 210128

Location

KCPD Traffic

AS IV Serial no: 097438
Version no: 532B

TEST RECORD 00373

Temp Date Time ^{a/} 210L

VOID: RPI
12 03/05/23 03:30

Subject Name

RPI

Subject I.D.

Operator Name, I.D.

Infranca 210128

Location

KCPD Traffic



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II
JORDAN INFRANCA**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/21/2021

NUMBER 210128

EXPIRES 6/21/2023

MS 880-8771 (8-19)

Laura Q. Day
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Bill Kuehl
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES LAB-4 (9-8-19)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

This instrument card is to be attached to operators as evidence of operator identity. It is not valid for the determination of the alcoholic content of breath from a subject at the instrument.

Operator: INFRANCA, JORDAN
Permit No: 210128
Date Issued: 6/21/2021 Date Expires: 6/21/2023



Airgas USA LLC (LAB)
3620 Barnard Street
St. Louis, Mo. 63103
Ph: (314) 533-8100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2051 Craig Road
St. Louis, Mo 63146

Test Date: 28-Nov-2021

Lot # AG132803 Model 108

Exp Date: 24-Nov-2023 Cyl. Type: 108 Component: Ethanol Nitrogen Certified Concentration: 0.082 ± 0.002 BAC (223 ppm)

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.3 ppm	EB0010559	268.3 ppm
EB0010285	208.0 ppm	EB0010585	208.3 ppm
EB0010584	103.6 ppm	EB0010562	104.2 ppm
EB0010651	52.12 ppm	EB0010573	52.31 ppm

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC434988	890.0 ppm	0058849	390.4 ppm
CC234903	255.0 ppm	0056682	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Intoximeters, Inc. (LAB)
Intoximeters, Inc. USA, LLC (LAB)
DN: cn=Quality Control

Approved for Release: *Rod Marsala*
Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07