

RECEIVED

By Tracy Crews at 1:38 pm, Jul 21, 2023



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097438	NAME OF AGENCY Kansas City Police Department	DATE OF INSPECTION 07-17-23
LOCATION OF INSTRUMENT (STREET AND CITY) 9701 MARION PARK DR, KANSAS CITY		TIME OF INSPECTION 0219

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER INTOXIMETERS LOT # AG132803 EXP. DATE 11/24/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1	.080	TEST 2	.080	TEST 3	.08
--------	-------------	--------	-------------	--------	------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	5	(0-.04)	0	(.05-.09)	0	(.10-.14)	3	(.15-.19)	4	(OVER .19)	0
----------	----------	---------	----------	-----------	----------	-----------	----------	-----------	----------	------------	----------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument meets all DOHSS standards and guidelines.

INSPECTING OFFICER

SIGNATURE
P.O. J. Infranca

PRINT NAME
P.O. Infranca 5670

TYPE II PERMIT NUMBER/EXPIRATION DATE
210128 06/21/2023

TELEPHONE NUMBER
() 816-382-5897

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 097438
Version no: 532B

TEST RECORD 00442 s/
Temp Date Time 210L

Air Blank: 07/17/23 02:13 .000
Subject Test: Man 21 07/17/23 02:13 .000

Subject Name
Test #1
Subject I.D.

Operator Name, I.D.
Infrans 230118
Location

KCPD Traffic

AS IV Serial no: 097438
Version no: 532B

TEST RECORD 00443 s/
Temp Date Time 210L

Air Blank: 07/17/23 02:15 .000
Subject Test: Man 21 07/17/23 02:15 .000

Subject Name
Test #2
Subject I.D.

Operator Name, I.D.
Infrans 230118
Location

KCPD Traffic

AS IV Serial no: 097438
Version no: 532B

TEST RECORD 00444 s/
Temp Date Time 210L

Air Blank: 07/17/23 02:17 .000
Subject Test: Man 22 07/17/23 02:17 .000

Subject Name
Test #3
Subject I.D.

Operator Name, I.D.
Infrans 230118
Location

KCPD Traffic

AS IV Serial no: 097438
Version no: 532B

TEST RECORD 00445 s/
Temp Date Time 210L

Air Blank: 07/17/23 02:19
Subject Test: Man 12 07/17/23 02:19

Subject Name
RFI
Subject I.D.

Operator Name, I.D.
Infrans 230118
Location

KCPD Traffic



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

JORDAN INFRANCA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 806.111 through 806.119 RSMo.

Mike Mazzoni
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DATE 6/7/2023
NUMBER 230118
EXPIRES 6/7/2025
MO 890-0771 (6-19)

David J. Nicholas
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LAB-4 (6-19)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD
The named contributor is authorized to operate an ethanol breath alcohol analyzer for the determination of the alcoholic content of breath from or expired air in Missouri.
Operator: INFRANCA, JORDAN
Permit No: 230118
Date Issued: 6/7/2023
Date Expires: 6/7/2025




Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 593-8100
Fax: (314) 593-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 29-Nov-2021

Lot # AG132803 Model 108

Exp Date 24-Nov-2023
Cyl. Type 108
Component Ethanol Nitrogen
Certified Concentration 0.062 ± 0.002 BRAC (223 ppm)

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	292.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010564	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC434668	800.0 ppm	0058649	390.1 ppm
CC234503	253.0 ppm	0058662	150.2 ppm

Analytical Method: NDIR

Dr. Rod Marsella, Inc. Quality Control
Laboratory for Standardization of Analytical
Data 11/29/2021 12:42

Approved for Release: *Rod Marsella*
Rod Marsella

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07