

AS IV Serial no: 102457
Version no: 532B

TEST RECORD 01055

Temp Date Time ^{s/} 210L

Air Blank:
04/13/23 01:40 .000
Calibration Check:
23 04/13/23 01:40 .102

Subject Name

Test # 1
Subject I.D.

Hochne, Mark D. 2078
Operator Name, I.D.

Location

AS IV Serial no: 102457
Version no: 532B

TEST RECORD 01056

Temp Date Time ^{s/} 210L

Air Blank:
04/13/23 01:42 .000
Calibration Check:
23 04/13/23 01:42 .102

Subject Name

Test # 2
Subject I.D.

Hochne, Mark D. 2078
Operator Name, I.D.

Location

AS IV Serial no: 102457
Version no: 532B

TEST RECORD 01057

Temp Date Time ^{s/} 210L

Air Blank:
04/13/23 01:44 .000
Calibration Check:
23 04/13/23 01:44 .101

Subject Name

Test # 3
Subject I.D.

Hochne, Mark D. 2078
Operator Name, I.D.

Location

AS IV Serial no: 102457
Version no: 532B

TEST RECORD 01058

Temp Date Time ^{s/} 210L

VOID: RFI
12 04/13/23 01:46

Subject Name

RFI check
Subject I.D.

Hochne, Mark D. 2078
Operator Name, I.D.

Location

AS IV Serial no: 102457
Version no: 532B

TEST RECORD 01059

Temp Date Time ^{s/} 210L

Air Blank:
04/13/23 01:47 .000
Subject Test: Auto
23 04/13/23 01:47 .000

Subject Name

Self Test
Subject I.D.

Hochne, Mark D. 2078
Operator Name, I.D.

Location



Airgas USA LLC (LAB)
 8628 Barnhart Street
 St. Louis, Mo. 63103
 PH: (314) 838-8400
 FAX: (314) 838-7300

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Prochiron, Inc.
 2041 Crump Road
 St. Louis, Mo 63140

Test Date: 11-Apr-2022

Lot # AG209701 Model 100

Exp Date: 7-Apr-2024
 Cyl Type: 100
 Component: Ethanol Nitrogen
 Certified Concentration: 0.100 ± 2% Std (272 ppm)

Certification Traceable to NIST, CRM and to CRM Ethanol Standard

CRM Serial No.	Concentration	CRM Serial No.	Concentration
MS0010601	341.8 ppm	MS0010602	392.8 ppm
MS0010670	280.8 ppm	MS0010603	291.9 ppm
MS0010606	204.0 ppm	MS0010604	184.2 ppm
MS0010601	108.7 ppm	MS0010605	52.91 ppm
MS0010601	52.22 ppm	MS0010606	

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727401	300.0 ppm	CC727402	300.9 ppm
CC727402	283.0 ppm	CC727403	180.9 ppm

Analytical Method: NDIR

Equally important to the quality of the product is the quality of the analysis. Airgas USA LLC (LAB) is ISO 9001:2015 certified.

Approved for Release: Robt. Marsala
 Robt Marsala

ISO 17025:2017 AZLA accredited. Certificate Number 3012.06
 ISO 17034:2016 AZLA accredited. Certificate Number 3002.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

MARK D HOEHNE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.080 through 577.041, RSMo and 508.111 through 508.119 RSMo.

DATE 6/14/2004

NUMBER 220158

EXPIRES 6/14/2004

MSD 400-0034 (6-03)

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Robert N. ...

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LSH-4 (03-03)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The person operating this instrument is responsible for ensuring accurate breath alcohol analysis and for the determination of the concentration of the breath alcohol sample as required by statute.

Operator: **MARK D HOEHNE**
 Permit No: **220158**
 Issue Date: **6/14/2004**

