

**RECEIVED**

By Tracy Crews at 8:03 am, Nov 09, 2023



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 102458 093.3578.252	NAME OF AGENCY Brookfield Police Dept.	DATE OF INSPECTION 11/08/2023
LOCATION OF INSTRUMENT (STREET AND CITY) 116 W. Brppks St. Brookfield, MO 64628		TIME OF INSPECTION 2:40 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Lab LOT # 23180 EXP. DATE 05/17/2025

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIM. SN SD 2763 SIM. NIST EXP DATE 01/27/2024

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1  0.099

TEST 2  0.099

TEST 3  0.098

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

This instrument meets Dept. of Health standards.

**INSPECTING OFFICER**

SIGNATURE

PRINT NAME  
Stuart Hughes

TYPE II PERMIT NUMBER/EXPIRATION DATE  
220175 06/29/2024

TELEPHONE NUMBER  
(660) 258-3385

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS 1U Serial no: 102458  
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 01061

Temp Date Time 210L %/

Air Blank: 11/08/23 14:40 .000  
Calibration Check: 27 11/08/23 14:40 .099

Subject Name  
S. Hughes  
Subject I.D.

Operator Name, I.D.

S. Hughes 220175

Location

BPD

AS 1U Serial no: 102458  
Version no: 532B

TEST RECORD 01062

Temp Date Time 210L %/

Air Blank: 11/08/23 14:45 .000  
Calibration Check: 28 11/08/23 14:45 .099

Subject Name  
S. Hughes  
Subject I.D.

Operator Name, I.D.

S. Hughes 220175

Location

BPD

AS 1U Serial no: 102458  
Version no: 532B

TEST RECORD 01063

Temp Date Time 210L %/

Air Blank: 11/08/23 14:46 .000  
Calibration Check: 28 11/08/23 14:46 .098

Subject Name  
S. Hughes  
Subject I.D.

Operator Name, I.D.

S. Hughes 220175

Location

BPD

AS 1U Serial no: 102458  
Version no: 532B

TEST RECORD 01064

Temp Date Time 210L %/

NOTE: RFI  
12 11/08/23 14:48

Subject Name  
RFI  
Subject I.D.

Operator Name, I.D.

S. Hughes 220175

Location

BPD

**GUTH LABORATORIES, INC.**

590 NORTH 57th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-684-6470

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23180** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1220%** (w/vol) ethyl alcohol. The expiration date for this lot number is **May 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L  $\pm$  3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN11172002 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

PERMIT  
TYPE II

STUART W. HUGHES

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repair and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/29/2022

NUMBER 220175

EXPIRES 6/29/2024

*Mike Mason*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*David E. Nicholson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 585-0771 (6-10)

LAB-4

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator HUGHES, STUART  
Permit No 220175  
Date Issued 6/29/2022 Date Expires 6/29/2024

