



AS IU Serial no: 102459  
Version no: 532B

TEST RECORD 00866

Temp Date Time 210L <sup>9/</sup>

Air Blank:  
01/24/24 21:37 .000  
Calibration Check:  
17 01/24/24 21:37 .104

Subject Name

Test 1

Subject I.D.

Operator Name, I.D.

Dickens 230251

Location

Zone 15 office

AS IU Serial no: 102459  
Version no: 532B

TEST RECORD 00867

Temp Date Time 210L <sup>9/</sup>

Air Blank:  
01/24/24 21:39 .000  
Calibration Check:  
18 01/24/24 21:39 .104

Subject Name

Test 2

Subject I.D.

Operator Name, I.D.

Dickens 230251

Location

Zone 15 office

AS IU Serial no: 102459  
Version no: 532B

TEST RECORD 00868

Temp Date Time 210L <sup>9/</sup>

Air Blank:  
01/24/24 21:41 .000  
Calibration Check:  
19 01/24/24 21:41 .104

Subject Name

Test 3

Subject I.D.

Operator Name, I.D.

Dickens 230251

Location

Zone 15 office

AS IU Serial no: 102459  
Version no: 532B

TEST RECORD 00870

Temp Date Time 210L <sup>9/</sup>

VOID: RFI  
12 01/24/24 21:44

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

Dickens 230251

Location

Zone 15 office



## GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23180** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1220%** (w/vol) ethyl alcohol. The expiration date for this lot number is **May 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L  $\pm$  3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN11172002 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*





STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**MARYLYN A. DICKENS**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/7/2023

NUMBER 230251

EXPIRES 11/7/2025

*Mike Massum*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*David J. Dickerson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** DICKENS, MARYLYN  
**Permit No** 230251  
**Date Issued** 11/7/2023 **Date Expires** 11/7/2025

