



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

RECEIVED  
 BY: Tracy D. Brown at 8:55 am, Mar 24

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

|   |  |                                  |
|---|--|----------------------------------|
| ALCO SENSOR IV SN<br>102461   | NAME OF AGENCY<br>Sullivan Police Department | DATE OF INSPECTION<br>03/23/2023 |
| LOCATION OF INSTRUMENT (STREET AND CITY)<br>106 Progress Drive Sullivan, MO 63080 |  | TIME OF INSPECTION<br>7:35 am    |

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION                       COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Labs                      LOT # 22310                      EXP. DATE 08/11/2024

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0                      SIM. SN SD 3322                      SIM. NIST EXP DATE 01/20/2024

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

|               |               |               |
|---------------|---------------|---------------|
| TEST 1 ← .096 | TEST 2 ← .096 | TEST 3 ← .097 |
|---------------|---------------|---------------|

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

|          |   |         |   |           |   |           |   |           |   |            |   |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|
| REFUSALS | 0 | (0-.04) | 1 | (.05-.09) | 0 | (.10-.14) | 0 | (.15-.19) | 0 | (OVER .19) | 0 |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument operating within Department of Health specifications.

**INSPECTING OFFICER**

|  |                                    |
|--|------------------------------------|
| SIGNATURE<br>  | PRINT NAME<br>Jason R. Stockton    |
| TYPE II PERMIT NUMBER/EXPIRATION DATE<br>210112/05-18-23 | TELEPHONE NUMBER<br>(573) 468-8001 |

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

LOG

AS IV Serial no: 102461  
Version no: 532B

TEST RECORD 01059

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
03/23/23 07:37 .000  
Calibration Check:  
24 03/23/23 07:37 .096

Subject Name

Subject I.D.

Operator Name, I.D.

Location

LOG

AS IV Serial no: 102461  
Version no: 532B

TEST RECORD 01059

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
03/23/23 07:39 .000  
Calibration Check:  
25 03/23/23 07:39 .096

Subject Name

Subject I.D.

Operator Name, I.D.

Location

LOG

AS IV Serial no: 102461  
Version no: 532B

TEST RECORD 01060

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
03/23/23 07:40 .000  
Calibration Check:  
25 03/23/23 07:40 .097

Subject Name

Subject I.D.

Operator Name, I.D.

Location

LOG

AS IV Serial no: 102461  
Version no: 532B

TEST RECORD 01061

Temp Date Time <sup>s/</sup> 210L

VOID: RPI  
12 03/23/23 07:42

Subject Name

Subject I.D.

Operator Name, I.D.

Location



## GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111- 4511 • TELEPHONE: 717-564-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 22310 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on August 16, 2022, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1205% (w/vol) ethyl alcohol. The expiration date for this lot number is August 11, 2024 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**JASON R. STOCKTON**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/18/2021

NUMBER 210112

EXPIRES 5/18/2023

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator STOCKTON, JASON  
 Permit No 210112  
 Date Issued 5/18/2021 Date Expires 5/18/2023