



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 9:49 pm, Dec 30, 2023

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 102473	NAME OF AGENCY GREENE COUNTY SHERIFF'S OFFICE	DATE OF INSPECTION 12/26/2023
LOCATION OF INSTRUMENT (STREET AND CITY) 5100 W DIVISION STREET SPRINGFIELD, MISSOURI		TIME OF INSPECTION 12:38

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER INTOXIMETERS _____ LOT # AG306807 _____ EXP. DATE 03/09/2025 _____

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIM. SN _____ SIM. NIST EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .100

TEST 2 ← .100

TEST 3 ← .099

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

PRINT NAME
 KYLE WINCHELL

TYPE II PERMIT NUMBER/EXPIRATION DATE
 220200 08/19/2024

TELEPHONE NUMBER
 (417) 868-4040

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 102473
Version no: 532B

TEST RECORD 02052

Temp Date Time ^{s/} 210L

Air Blank:
12/26/23 12:38 .000
Calibration Check:
20 12/26/23 12:38 .100

Subject Name

Test 1

Subject I.D.

Operator Name, I.D.

JL Judd

Location

AS IV Serial no: 102473
Version no: 532B

TEST RECORD 02053

Temp Date Time ^{s/} 210L

Air Blank:
12/26/23 12:40 .000
Calibration Check:
20 12/26/23 12:40 .100

Subject Name

Test 2

Subject I.D.

Operator Name, I.D.

JL Judd

Location

AS IV Serial no: 102473
Version no: 532B

TEST RECORD 02054

Temp Date Time ^{s/} 210L

Air Blank:
12/26/23 12:42 .000
Calibration Check:
21 12/26/23 12:42 .099

Subject Name

Test 3

Subject I.D.

Operator Name, I.D.

JL Judd

Location

AS IV Serial no: 102473
Version no: 532B

TEST RECORD 02055

Temp Date Time ^{s/} 210L

VOID: RFI
12 12/26/23 12:44

Subject Name

Test 4 / RFI

Subject I.D.

Operator Name, I.D.

JL Judd

Location

AS IV Serial no: 102473
Version no: 532B

TEST RECORD 02056

Temp Date Time ^{s/} 210L

Air Blank:
12/26/23 12:45 .000
Subject Test: Auto
22 12/26/23 12:45 .000

Subject Name

Test 5 / Sober

Subject I.D.

Operator Name, I.D.

JL Judd

Location



Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 13-Mar-2023

Lot # AG306807 **Model** 108

Exp Date	Cyl. Type	Component	Certified Concentration
9-Mar-2025	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm)

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by:Quality Control
Reason:Dry gas standard certification of analysis
Location:Airgas USA LLC (Lab)
Date:03.16.2023 13:02

Approved for Release: _____

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
KYLE R. WINCHELL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/19/2022

NUMBER 220200

EXPIRES 8/19/2024

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator WINCHELL, KYLE
Permit No 220200
Date Issued 8/19/2022 Date Expires 8/19/2024

