



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

RECEIVED

By Brianna Medrano at 8:08 am, Apr 28, 2023

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 104641	NAME OF AGENCY O'Fallon PD	DATE OF INSPECTION 4-18-2023
LOCATION OF INSTRUMENT (STREET AND CITY) 1019 Bryan Road, O'Fallon, MO		TIME OF INSPECTION 1855

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters</u>	LOT # <u>AG115405</u> EXP. DATE <u>06-03-2023</u>
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C)	SIM. SN _____ SIM. NIST EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← <u>.079</u>	TEST 2 ← <u>.079</u>	TEST 3 ← <u>.079</u>
----------------------	----------------------	----------------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <u>0</u>	(0-.04) <u>0</u>	(.05-.09) <u>0</u>	(.10-.14) <u>0</u>	(.15-.19) <u>0</u>	(OVER .19) <u>0</u>
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

N/A

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME Chris Sinnokrak
TYPE II PERMIT NUMBER/EXPIRATION DATE 220271 / 12-14-2024	TELEPHONE NUMBER (636)240-3200

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 104641
Version no: 532B

TEST RECORD 01340

Temp Date Time 210L ^{s/}

Air Blank:
04/18/23 18:57 .000
Calibration Check:
21 04/18/23 18:57 .079

Subject Name

TEST 1

Subject I.D.

Operator Name, I.D.

SINACKRAK 334

Location

AS IV Serial no: 104641
Version no: 532B

TEST RECORD 01341

Temp Date Time 210L ^{s/}

Air Blank:
04/18/23 18:58 .000
Calibration Check:
21 04/18/23 18:58 .079

Subject Name

TEST 2

Subject I.D.

Operator Name, I.D.

SINACKRAK 334

Location

AS IV Serial no: 104641
Version no: 532B

TEST RECORD 01342

Temp Date Time 210L ^{s/}

Air Blank:
04/18/23 19:00 .000
Calibration Check:
22 04/18/23 19:00 .079

Subject Name

TEST 3

Subject I.D.

Operator Name, I.D.

SINACKRAK 334

Location

AS IV Serial no: 104641
Version no: 532B

TEST RECORD 01343

Temp Date Time 210L ^{s/}

VOID: RFI
12 04/18/23 19:02

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 104641
Version no: 532B

TEST RECORD 01344

Temp Date Time 210L ^{s/}

Air Blank:
04/18/23 19:03 .000
Subject Test: Auto
24 04/18/23 19:03 .000

Subject Name

SOBET TEST

Subject I.D.

Operator Name, I.D.

Location



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
CHRIS SINNOKRAK

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/14/2022

NUMBER 220271

EXPIRES 12/14/2024

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nielson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator SINNOKRAK, CHRIS
Permit No 220271
Date Issued 12/14/2022 **Date Expires** 12/14/2024





Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 7-Jun-2021

Lot # AG115405 Model 108cacd

Exp. Date

3-Jun-2023

Cyl. Type

108

Component

Ethanol
Nitrogen

Certified Concentration

0.080 ± 0.002 BrAC (218 ppm)
Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.

EB0010581

EB0010570

EB0010285

EB0010561

EB0010681

Concentration

392.1 ppm

259.8 ppm

208.0 ppm

103.6 ppm

52.12 ppm

RGM Serial No.

EB0010603

EB0010559

EB0010595

EB0010562

EB0010579

Concentration

393.0 ppm

258.2 ppm

208.3 ppm

104.2 ppm

52.81 ppm

CRM Serial No.

CC434668

CC234503

Concentration

800.0 ppm

253.0 ppm

CRM Serial No.

0056649

0056662

Concentration

390.1 ppm

150.2 ppm

Analytical Method:

NDIR

Digitally signed by Quality Control
Date: 2021.06.08 13:46:01 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06

ISO 17034:2016 A2LA accredited. Certificate Number 3082.07