



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 BY: TERRY CHENEY AT 12:55pm On 10/16/23

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| | | |
|-----------------------------|---|----------------------------------|
| ALCO SENSOR IV SN 105443 | NAME OF AGENCY Raymore Police Department | DATE OF INSPECTION 10/14/2023 |
|-----------------------------|---|----------------------------------|

| | |
|--|-------------------------------|
| LOCATION OF INSTRUMENT (STREET AND CITY) 100 Municipal Circle Raymore, Missouri 64083 | TIME OF INSPECTION 2:02 pm |
|--|-------------------------------|

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories, Inc. LOT # 22310 EXP. DATE 08/11/2024

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIM. SN SD2256 SIM. NIST EXP DATE 10/05/2024

CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

| | | |
|---------------|---------------|---------------|
| TEST 1 ☛ .099 | TEST 2 ☛ .099 | TEST 3 ☛ .099 |
|---------------|---------------|---------------|

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

| | | | | | | | | | | | |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|
| REFUSALS | 0 | (0-.04) | 0 | (.05-.09) | 0 | (.10-.14) | 0 | (.15-.19) | 0 | (OVER .19) | 0 |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

| | |
|--|-------------------------------|
| SIGNATURE <i>Joshua Giaccone #124</i> | PRINT NAME Joshua Giaccone |
|--|-------------------------------|

| | |
|--|------------------------------------|
| TYPE II PERMIT NUMBER/EXPIRATION DATE 220105 03/29/2024 | TELEPHONE NUMBER (816) 331-0530 |
|--|------------------------------------|

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 105443
Version no: 532B

TEST RECORD 01365

Temp Date Time 210L^{g/}

Air Blank:
10/14/23 14:15 .000
Calibration Check:
23 10/14/23 14:15 .099

Subject Name

Maintenance

Subject I.D. #220105

J. Loiacono #924

Operator Name, I.D.

Palmore P.D.

Location

AS IV Serial no: 105443
Version no: 532B

TEST RECORD 01366

Temp Date Time 210L^{g/}

Air Blank:
10/14/23 14:17 .000
Calibration Check:
23 10/14/23 14:17 .099

Subject Name

Maintenance

Subject I.D. #220105

J. Loiacono #924

Operator Name, I.D.

Palmore P.D.

Location

AS IV Serial no: 105443
Version no: 532B

TEST RECORD 01367

Temp Date Time 210L^{g/}

Air Blank:
10/14/23 14:18 .000
Calibration Check:
24 10/14/23 14:18 .099

Subject Name

Maintenance

Subject I.D. #220105

J. Loiacono #924

Operator Name, I.D.

Palmore P.D.

Location

AS IV Serial no: 105443
Version no: 532B

TEST RECORD 01368

Temp Date Time 210L^{g/}

VOID: RFI
12 10/14/23 14:20

Subject Name

Maintenance

Subject I.D. #220105

J. Loiacono #924

Operator Name, I.D.

Palmore P.D.

Location



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **22310** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **August 16, 2022**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1205%** (w/vol) ethyl alcohol. The expiration date for this lot number is **August 11, 2024** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

*Testing was conducted using Cerilliant Reference Standard lot number **FN03052002** whose values are traceable to NIST.*

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



Paula Nickelson
Acting Director

Michael L. Parson
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: SD2256 Manufacturer: Guth
Model Number: 10-4D
Agency: RAYMORE PD
Agency Address: 100 MUNICIPAL CIRCLE, RAYMORE, MO 64083

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00690 Bias: 0.00
Uncertainty: 0.02
Date of Certification: 10/24/2022 Date of Expiration: 10/24/2023

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

Table with 3 columns: Simulator Average (34.00), NIST Average (33.98), Combined Uncertainty (.04)

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 10/5/2023
Certification Expiration: 10/5/2024
Simulator testing technician: R. SCHILDKNECHT

Notes on Condition: none
Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO
Certification No: SD2256_1052023

X Brianna Medrano (signature)

DHSS BAP Scientist Approving



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
JOSHUA B. GIACONE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/29/2022

NUMBER 220105

EXPIRES 3/29/2024

Laura P. Wag

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Douglas J. Nickelson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator GIACONE, JOSHUA
Permit No 220105
Date Issued 3/29/2022 Date Expires 3/29/2024

