

RECEIVED

By Tracy Crews at 7:32 am, Oct 16, 2023



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 105448 / Printer 09B.3591.016	NAME OF AGENCY Clinton P.D.	DATE OF INSPECTION 10/16/2023
LOCATION OF INSTRUMENT (STREET AND CITY) 101 East Ohio Street, Clinton, MO 64735		TIME OF INSPECTION 05:01 a.m.

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Guth Laboratories, Inc.</u> LOT # <u>22310</u> EXP. DATE <u>8/11/2024</u>	
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>34.0</u> SIM. SN <u>SD3509</u> SIM. NIST EXP DATE <u>08/08/2024</u>	

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .100	TEST 2 ← .100	TEST 3 ← .098
----------------------	----------------------	----------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	----------	---------	----------	-----------	----------	-----------	----------	-----------	----------	------------	----------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

- Added print paper

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME Michael Nelson
TYPE II PERMIT NUMBER/EXPIRATION DATE 220186 / 07-22-2024	TELEPHONE NUMBER (660) 885-2679

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 105448
Version no: 532B

TEST RECORD 01765

Temp Date Time ^{g/} 210L

Air Blank:
10/16/23 05:01 .000
Calibration Check:
23 10/16/23 05:01 .100

Subject Name
TEST #1

Subject I.D.
N/A

Operator Name, I.D.
M. Nelson #220186

Location
Clinton P.D.
101 E. Ohio St.
Clinton, MO

AS IV Serial no: 105448
Version no: 532B

TEST RECORD 01767

Temp Date Time ^{g/} 210L

Air Blank:
10/16/23 05:08 .000
Calibration Check:
24 10/16/23 05:08 .100

Subject Name
TEST #2

Subject I.D.
N/A

Operator Name, I.D.
M. Nelson #220186

Location
Clinton P.D.
101 E. Ohio St.
Clinton, MO

AS IV Serial no: 105448
Version no: 532B

TEST RECORD 01768

Temp Date Time ^{g/} 210L

Air Blank:
10/16/23 05:11 .000
Calibration Check:
24 10/16/23 05:11 .098

Subject Name
TEST #3

Subject I.D.
N/A

Operator Name, I.D.
M. Nelson #220186

Location
Clinton P.D.
101 E. Ohio St.
Clinton, MO

AS IV Serial no: 105448
Version no: 532B

TEST RECORD 01769

Temp Date Time ^{g/} 210L

VOID: RFI
12 10/16/23 05:13

Subject Name
RFI

Subject I.D.
N/A

Operator Name, I.D.
M. Nelson #220186

Location
Clinton P.D.
101 E. Ohio St.
Clinton, MO

AS IV Serial no: 105448
Version no: 532B

TEST RECORD 01770

Temp Date Time ^{g/} 210L

Air Blank:
10/16/23 05:17 .000
Subject Test: Auto
25 10/16/23 05:17 .000

Subject Name
SELF TEST

Subject I.D.
N/A

Operator Name, I.D.
M. Nelson #220186

Location
Clinton P.D.
101 E. Ohio St.
Clinton, MO



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 22310 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on August 16, 2022, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1205% (w/vol) ethyl alcohol. The expiration date for this lot number is August 11, 2024 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
MICHAEL S. NELSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/22/2022

NUMBER 220186

EXPIRES 7/22/2024

MO 580-0771 (6-10)

Mike Masman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nielson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator NELSON, MICHAEL
Permit No 220186
Date Issued 7/22/2022 **Date Expires** 7/22/2024

