



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 10541	NAME OF AGENCY Breceknridge Hills Police Department	DATE OF INSPECTION 07/05/2023
LOCATION OF INSTRUMENT (STREET AND CITY) 9623 St. Charles Rock Road, St. Louis MO 63114		TIME OF INSPECTION 4:08 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER GUTH LABS LOT # 22430 EXP. DATE 11/30/2024
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIM. SN MP653 SIM. NIST EXP DATE 05/04/2024

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .104	TEST 2 .105	TEST 3 .106
--------------	--------------	--------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	1	(.10-.14)	0	(.15-.19)	2	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Randy Von Cloedt
TYPE II PERMIT NUMBER/EXPIRATION DATE 220142 5/11/24	TELEPHONE NUMBER (314) 426-1214

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS 10 Serial no: 105451
Version no: 532B

TEST RECORD 01709

Temp Date Time 9/
210L

Air Blank:
07/05/23 16:23 .000
Calibration Check:
22 07/05/23 16:23 .104

Subject Name

Maintenance

Subject I.D.

N/A

Operator Name, I.D.

VanClant 207

Location

9623 MOWO

St. Louis MO

63119

AS 10 Serial no: 105451
Version no: 532B

TEST RECORD 01710

Temp Date Time 9/
210L

Air Blank:
07/05/23 16:25 .000
Calibration Check:
22 07/05/23 16:25 .105

Subject Name

Maint.

Subject I.D.

N/A

Operator Name, I.D.

VanClant 207

Location

9623 MOWO

St. Louis MO 63114

63119

AS 10 Serial no: 105451
Version no: 532B

TEST RECORD 01714

Temp Date Time 9/
210L

Air Blank:
07/05/23 16:43 .000
Calibration Check:
25 07/05/23 16:43 .104

Subject Name

Maint.

Subject I.D.

N/A

Operator Name, I.D.

VanClant 207

Location

9623 MOWO

St. Louis MO 63114

63119

AS 10 Serial no: 105451
Version no: 532B

TEST RECORD 01712

Temp Date Time 9/
210L

WIDE RFI
12 07/05/23 16:32

Subject Name

Maint.

Subject I.D.

N/A

Operator Name, I.D.

VanClant 207

Location

9623 MOWO

St. Louis MO 63114

63119



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **22430** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 1, 2022**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1216%** (w/vol) ethyl alcohol. The expiration date for this lot number is **November 30, 2024** at 11:59 PM.

When used in a calibrated Simulator, operating at **34°C +/- .2°C**, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

RANDY J. VON CLOEDT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/11/2022

NUMBER 220142

EXPIRES 5/11/2024

Laura A. Neag

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David F. Nielsen

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (5-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator VON CLOEDT, RANDY
 Permit No 220142
 Date Issued 5/11/2022 Date Expires 5/11/2024

