



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
By Tracy Crews at 9:10 am, Jul 06, 2023

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 107999	NAME OF AGENCY Clay County Sheriff's Office	DATE OF INSPECTION 07/04/2023
-----------------------------	--	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 12 S. Water St., Liberty, MO 64068	TIME OF INSPECTION 12:02 am
--	--------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters</u>	LOT # <u>AG221502</u> EXP. DATE <u>08/03/2024</u>
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C)	SIM. SN _____ SIM. NIST EXP DATE _____

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .096	TEST 2 ← .097	TEST 3 ← .097
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	2	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME Tom Butkovich
TYPE II PERMIT NUMBER/EXPIRATION DATE 220065 - 02/25/24	TELEPHONE NUMBER (816) 407-3702

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 107999
Version no: 532B

TEST RECORD 00068

Temp Date Time ^{s/} 210L

Air Blank:
07/04/23 00:04 .000
Calibration Check:
26 07/04/23 00:04 .096

Subject Name

Test #1

Subject I.D.

Butkovich 8237

Operator Name, I.D.

Location

CCSO

AS IV Serial no: 107999
Version no: 532B

TEST RECORD 00069

Temp Date Time ^{s/} 210L

Air Blank:
07/04/23 00:05 .000
Calibration Check:
26 07/04/23 00:05 .097

Subject Name

Test #2

Subject I.D.

Butkovich 8237

Operator Name, I.D.

Location

CCSO

AS IV Serial no: 107999
Version no: 532B

TEST RECORD 00070

Temp Date Time ^{s/} 210L

Air Blank:
07/04/23 00:07 .000
Calibration Check:
27 07/04/23 00:07 .097

Subject Name

Test #3

Subject I.D.

Butkovich 8237

Operator Name, I.D.

CCSO

Location

AS IV Serial no: 107999
Version no: 532B

TEST RECORD 00071

Temp Date Time ^{s/} 210L

VOID: RFI
12 07/04/23 00:09

Subject Name

RFI Test

Subject I.D.

Butkovich 8237

Operator Name, I.D.

Location

CCSO



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
THOMAS F. BUTKOVICH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/5/2023

NUMBER 230113

EXPIRES 6/5/2025

Mike Masson

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nielsen

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BUTKOVICH, THOMAS
 Permit No 230113
 Date Issued 6/5/2023 Date Expires 6/5/2025





Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 4-Aug-2022

Lot # AG221502 Model 108

Exp Date	Cyl. Type	Component	Certified Concentration
3-Aug-2024	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm)


Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by:Quality Control
 Reason:Dry gas standard certification of analysis
 Location:Airgas USA LLC (Lab)
 Date:08.04.2022 15:01

Approved for Release: 
 Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07