

**RECEIVED**

By Tracy Crews at 2:35 pm, Apr 05, 2023

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 108271	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 04/03/2023
LOCATION OF INSTRUMENT (STREET AND CITY) 3131 E Kearney, Springfield, Missouri 65803		TIME OF INSPECTION 09:41

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☒ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)☒ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)☒ PRINTER WORKING PROPERLY☒ TIME AND DATE DISPLAYING PROPERLY**BREATH ALCOHOL ACCURACY STANDARDS**☒ SIMULATOR SOLUTION ☐ COMPRESSED ETHANOL-GAS MIXTURE☒ STANDARD SUPPLIER REPCO LOT # 21001 EXP. DATE 06/16/2023☒ SIMULATOR TEMPERATURE (34°C ± 0.2°C) 33.99 SIM. SN MP2418 SIM. NIST EXP DATE 12/12/2023☒ CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .100

TEST 2 ← .100

TEST 3 ← .100

☒ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19) <u>1</u>	(OVER .19)
----------	---------	-----------	-----------	--------------------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**changed time +1 hour +1minute**

**INSPECTING OFFICER**

SIGNATURE <td>PRINT NAME D W Henley #727</td>	PRINT NAME D W Henley #727
TYPE II PERMIT NUMBER/EXPIRATION DATE 220190 8/3/2024	TELEPHONE NUMBER ( 417 ) 895 - 6868

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
by mail, fax, or email.

AS IV Serial no: 108271  
Version no: 532B

TEST RECORD 00698

Temp Date Time 210L

Air Blank: 04/03/23 09:45 .000

Calibration Check: 20 04/03/23 09:45 .100

Subject Name

TEST 1

Subject I.D.

Operator Name, I.D.

D W HENLEY #727

Location

3131 E KEARNEY

SPRINGFIELD, MO

DJ #727

AS IV Serial no: 108271  
Version no: 532B

TEST RECORD 00699

Temp Date Time 210L

Air Blank: 04/03/23 09:47 .000

Calibration Check: 20 04/03/23 09:47 .100

Subject Name

TEST 2

Subject I.D.

Operator Name, I.D.

D W HENLEY #727

Location

3131 E KEARNEY

SPRINGFIELD, MO

DJ #727

AS IV Serial no: 108271  
Version no: 532B

TEST RECORD 00700

Temp Date Time 210L

Air Blank: 04/03/23 09:48 .000

Calibration Check: 21 04/03/23 09:48 .100

Subject Name

TEST 3

Subject I.D.

Operator Name, I.D.

D W HENLEY #727

Location

3131 E KEARNEY

SPRINGFIELD, MO

DJ #727

AS IV Serial no: 108271  
Version no: 532B

TEST RECORD 00701

Temp Date Time 210L

VOID: RFI

12 04/03/23 09:50

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

D W HENLEY #727

Location

3131 E KEARNEY

SPRINGFIELD, MO

DJ #727

RepCo Marketing Co  
3101-188 Stony Brook Drive  
Raleigh, NC 27604  
919-876-5480

## CERTIFICATE OF ANALYSIS

**MANUFACTURER AND SUPPLIER:** RepCo Marketing Co.

**LOT NUMBER:** 21001

**EXPIRATION DATE:** June 16, 2023 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 21001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1222 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 17, 2021 The expiration date for this lot number is June 16, 2023 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

  
Alma Palmer, Operations Manager  
RepCo Marketing Co.



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**DAVID W. HENLEY, JR.**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/3/2022

NUMBER 220190

EXPIRES 8/3/2024

*Mike Mason*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*David J. Nicholson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HENLEY, JR., DAVID

Permit No 220190

Date issued 8/3/2022 Date Expires 8/3/2024

