

AS IV Serial no: 108388
Version no: 532B

TEST RECORD 00812

Temp Date Time ^{s/} 210L

Air Blank:
05/29/23 10:34 .000
Calibration Check:
19 05/29/23 10:34 .105

Subject Name

Test #1

Subject I.D.

N/A

Operator Name, I.D.

P.D. Toste 4553

Location

MODOT TMC

AS IV Serial no: 108388
Version no: 532B

TEST RECORD 00813

Temp Date Time ^{s/} 210L

Air Blank:
05/29/23 10:36 .000
Calibration Check:
20 05/29/23 10:36 .105

Subject Name

Test #2

Subject I.D.

N/A

Operator Name, I.D.

P.D. Toste 4553

Location

MODOT TMC

AS IV Serial no: 108388
Version no: 532B

TEST RECORD 00814

Temp Date Time ^{s/} 210L

Air Blank:
05/29/23 10:37 .000
Calibration Check:
20 05/29/23 10:37 .105

Subject Name

Test #3

Subject I.D.

N/A

Operator Name, I.D.

P.D. Toste 4553

Location

MODOT TMC

AS IV Serial no: 108388
Version no: 532B

TEST RECORD 00815

Temp Date Time ^{s/} 210L

VOID: RFI
12 05/29/23 10:39

Subject Name

RFI

Subject I.D.

N/A

Operator Name, I.D.

P.D. Toste 4553

Location

MODOT TMC



Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 19-Feb-2022

Lot # AG204801 **Model** 108

Exp Date	Cyl. Type	Component	Certified Concentration
17-Feb-2024	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm)

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by: Quality Control
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)
Date: 02.22.2022 19:37

Approved for Release: _____

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
ROBERT TOSIE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/24/2022

NUMBER 220173

EXPIRES 6/24/2024

Mike Massman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David J. Nielson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES