



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

By Tracy Crews at 9:23 am, Jun 05, 2023

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 110743	NAME OF AGENCY Saint Joseph Police Department	DATE OF INSPECTION 06/02/2023
LOCATION OF INSTRUMENT (STREET AND CITY) 501 Faraon Street, Saint Joseph MO. 64501		TIME OF INSPECTION 6:37 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG304601 EXP. DATE 02/15/2025

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 → 0.101

TEST 2 → 0.100

TEST 3 → .100

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS

(0-.04)

(.05-.09)

(.10-.14)

(.15-.19)

(OVER .19)

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

PRINT NAME

John L. Foster

TYPE II PERMIT NUMBER/EXPIRATION DATE

210197 Exp-09/09/2023

TELEPHONE NUMBER

(816) 596-8206

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

SAINT JOSEPH POLICE DEPARTMENT MONTHLY MAINTENANCE REPORT ASIV- W/PRINTER

<p>AS IV Serial no: 110743 Version no: 532B TEST RECORD 01967</p> <p>Temp Date Time 210L 9/</p> <p>VOID: RFI 12 06/02/23 06:37</p> <p>Subject Name <u>Matthew Test</u></p> <p>Subject I.D. <u>Foster JLn 28977</u></p> <p>Operator Name, I.D. <u>SO1 Fortaast</u></p> <p>Location <u>St Joe Mo 64501</u></p> <p><u>LEC</u></p>	<p>AS IV Serial no: 110743 Version no: 532B TEST RECORD 01968</p> <p>Temp Date Time 210L 9/</p> <p>Air Blank: 06/02/23 06:42 .000 Calibration Check: 21 06/02/23 06:42 .101</p> <p>Subject Name <u>Matthew Test</u></p> <p>Subject I.D. <u>Foster JLn 28977</u></p> <p>Operator Name, I.D. <u>SO Forta SE</u></p> <p>Location <u>St Joseph Mo 64501</u></p> <p><u>LEC</u></p>	<p>AS IV Serial no: 110743 Version no: 532B TEST RECORD 01969</p> <p>Temp Date Time 210L 9/</p> <p>Air Blank: 06/02/23 06:44 .000 Calibration Check: 22 06/02/23 06:44 .100</p> <p>Subject Name <u>Matthew Test</u></p> <p>Subject I.D. <u>Foster JLn 28977</u></p> <p>Operator Name, I.D. <u>SO Forta SE</u></p> <p>Location <u>St Joseph Mo</u></p> <p><u>LEC</u></p>	<p>AS IV Serial no: 110743 Version no: 532B TEST RECORD 01970</p> <p>Temp Date Time 210L 9/</p> <p>Air Blank: 06/02/23 06:46 .000 Calibration Check: 23 06/02/23 06:46 .100</p> <p>Subject Name <u>Matthew Test</u></p> <p>Subject I.D. <u>Foster JLn 28977</u></p> <p>Operator Name, I.D. <u>SO Forta ST</u></p> <p>Location <u>St Joseph Mo 64501</u></p> <p><u>LEC</u></p>
--	---	---	---



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
JOHN L. FOSTER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/9/2021

NUMBER 210197

EXPIRES 9/9/2023

Laura P. Way

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Donald A. Kamm

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

MO 580-0771 (6-10)

 **STATE OF MISSOURI**
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator FOSTER, JOHN
Permit No 210197
Date Issued 9/9/2021 Date Expires 9/9/2023

