

RECEIVED

By Tracy Crews at 9:02 am, Mar 06, 2023



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN: 111320 NAME OF AGENCY: St. Louis County Police Department DATE OF INSPECTION: 03/05/2023

LOCATION OF INSTRUMENT (STREET AND CITY): 14301 SOUTH OUTER FOURTY MODOT TMC TIME OF INSPECTION: 8:27 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [X] DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
[X] TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
[X] PRINTER WORKING PROPERLY
[X] TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- [] SIMULATOR SOLUTION [X] COMPRESSED ETHANOL-GAS MIXTURE
[X] STANDARD SUPPLIER Intoximeters LOT # AG204801 EXP. DATE 02/17/2024
[] SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

- [X] CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
[X] 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
[] 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
[] 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .101 TEST 2 .100 TEST 3 .101

- [X] RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0 (0-.04) 0 (.05-.09) 0 (.10-.14) 1 (.15-.19) 1 (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE: [Signature] 4553 PRINT NAME: PO Tosie DSN 4553

TYPE II PERMIT NUMBER/EXPIRATION DATE: 220173 / 06/24/2024 TELEPHONE NUMBER: (636) 529-8210

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111320
Version no: 532B

TEST RECORD 00558

Temp Date Time ^{a/} 210L

Air Blank:
03/05/23 08:27 .000
Calibration Check:
21 03/05/23 08:27 .101

Subject Name

Test #1

Subject I.D.

N/A

Operator Name, I.D.

P.O. Tosie 4553

Location

MODOT TMC

AS IV Serial no: 111320
Version no: 532B

TEST RECORD 00559

Temp Date Time ^{a/} 210L

Air Blank:
03/05/23 08:30 .000
Calibration Check:
22 03/05/23 08:30 .100

Subject Name

Test #2

Subject I.D.

N/A

Operator Name, I.D.

P.O. Tosie 4553

Location

MODOT TMC

AS IV Serial no: 111320
Version no: 532B

TEST RECORD 00560

Temp Date Time ^{a/} 210L

Air Blank:
03/05/23 08:31 .000
Calibration Check:
22 03/05/23 08:31 .101

Subject Name

Test #3

Subject I.D.

N/A

Operator Name, I.D.

P.O. Tosie 4553

Location

MODOT TMC

AS IV Serial no: 111320
Version no: 532B

TEST RECORD 00561

Temp Date Time ^{a/} 210L

VOID: REI
12 03/05/23 08:33

Subject Name

REI

Subject I.D.

N/A

Operator Name, I.D.

P.O. Tosie 4553

Location

MODOT TMC



Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 19-Feb-2022

Lot # AG204801 **Model** 108

Exp Date 17-Feb-2024	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0.100 ± 2% BrAC (272 ppm)
--------------------------------	-------------------------	---	---

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by:Quality Control
Reason:Dry gas standard certification of analysis
Location:Airgas USA LLC (Lab)
Date:02.22.2022 19:37

Approved for Release: Rod Marsala
Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
ROBERT TOSIE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/24/2022

NUMBER 220173

EXPIRES 6/24/2024

Mike Massimo

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David F. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES