



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| | | |
|---------------------------------------|--|----------------------------------|
| ALCO SENSOR IV SN 111327 Lawson PD | NAME OF AGENCY Lawson Police Department | DATE OF INSPECTION 12/29/2023 |
|---------------------------------------|--|----------------------------------|

| | |
|--|-------------------------------|
| LOCATION OF INSTRUMENT (STREET AND CITY) 104 W 3rd Street, Lawson, MO | TIME OF INSPECTION 4:01 am |
|--|-------------------------------|

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

| | |
|--|--|
| <input checked="" type="checkbox"/> SIMULATOR SOLUTION | <input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE |
| <input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Guth Laboratories</u> | LOT # <u>22430</u> EXP. DATE <u>11/02/2024</u> |
| <input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>34.0</u> | SIM. SN <u>DR5392</u> SIM. NIST EXP DATE <u>03/22/2024</u> |

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

| | | |
|---------------|---------------|---------------|
| TEST 1 ➡ .100 | TEST 2 ➡ .099 | TEST 3 ➡ .099 |
|---------------|---------------|---------------|

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

| | | | | | | | | | | | |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|
| REFUSALS | 0 | (0-.04) | 0 | (.05-.09) | 0 | (.10-.14) | 0 | (.15-.19) | 0 | (OVER .19) | 0 |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Corrected time, two minutes slow.

INSPECTING OFFICER

| | |
|--|------------------------------------|
| SIGNATURE <i>Chief Bruce Summa</i> | PRINT NAME Chief Bruce Summa |
| TYPE II PERMIT NUMBER/EXPIRATION DATE 220062 02/19/2024 | TELEPHONE NUMBER (816) 580-7210 |

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111327
Version no: 532B

TEST RECORD 00885

Temp Date Time ^{s/} 210L

Air Blank:
12/29/23 04:01 .000
Calibration Check:
26 12/29/23 04:01 .100

Subject Name

Monthly Maint

Subject I.D.

Chief Bruce Summa

Operator Name, I.D.

220062

Location

Lawson PD

Test 1

AS IV Serial no: 111327
Version no: 532B

TEST RECORD 00886

Temp Date Time ^{s/} 210L

Air Blank:
12/29/23 04:02 .000
Calibration Check:
26 12/29/23 04:02 .099

Subject Name

Monthly Maint

Subject I.D.

Chief Bruce Summa

Operator Name, I.D.

220062

Location

Lawson PD

Test 2

AS IV Serial no: 111327
Version no: 532B

TEST RECORD 00887

Temp Date Time ^{s/} 210L

Air Blank:
12/29/23 04:04 .000
Calibration Check:
26 12/29/23 04:04 .099

Subject Name

Monthly Maint

Subject I.D.

Chief Bruce Summa

Operator Name, I.D.

220062

Location

Lawson PD

Test 3

AS IV Serial no: 111327
Version no: 532B

TEST RECORD 00888

Temp Date Time ^{s/} 210L

VOID: RFI
12 12/29/23 04:05

Subject Name

Monthly Maint

Subject I.D.

Chief Bruce Summa

Operator Name, I.D.

220062

Location

Lawson PD

RFI Test



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **22430** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 1, 2022**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1216%** (w/vol) ethyl alcohol. The expiration date for this lot number is **November 30, 2024** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
BRUCE SUMMA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/19/2022

NUMBER 220062

EXPIRES 2/19/2024

Laura Q. Day

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Bruce Summa

, acting director
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R5-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator SUMMA, BRUCE
Permit No 220062
Date Issued 2/19/2022 **Date Expires** 2/19/2024

