

RECEIVED

By Tracy Crews at 8:22 am, Jun 23, 2023



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN **111633** NAME OF AGENCY **Kansas City Police Department** DATE OF INSPECTION **06/14/2023**

LOCATION OF INSTRUMENT (STREET AND CITY) **9701 MARION PARK DR, KANSAS CITY** TIME OF INSPECTION **0026**

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER **INTOXIMETERS** LOT # **AG132803** EXP. DATE **11/24/2023**

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 **.080** TEST 2 **.080** TEST 3 **.080**

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS **1** (0-.04) **0** (.05-.09) **1** (.10-.14) **2** (.15-.19) **2** (OVER .19) **5**

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument meets all DOHSS standards and guidelines.

INSPECTING OFFICER

SIGNATURE

PRINT NAME **Wade Robinson**

TYPE II PERMIT NUMBER/EXPIRATION DATE **210266 - 11/18/2023**

TELEPHONE NUMBER **816-482-8141**

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111633
Version no: 532B

TEST RECORD 00470

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
06/14/23 00:26 .000
Calibration Check:
24 06/14/23 00:26 .000

Subject Name

Test 1

Subject I.D.

Operator Name, I.D.

Robinson #210266
Location

AS IV Serial no: 111633
Version no: 532B

TEST RECORD 00471

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
06/14/23 00:27 .000
Calibration Check:
25 06/14/23 00:27 .000

Subject Name

Test 2

Subject I.D.

Operator Name, I.D.

Robinson #210266
Location

AS IV Serial no: 111633
Version no: 532B

TEST RECORD 00472

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
06/14/23 00:29 .000
Calibration Check:
25 06/14/23 00:29 .000

Subject Name

Test 3

Subject I.D.

Operator Name, I.D.

Robinson #210266
Location

AS IV Serial no: 111633
Version no: 532B

TEST RECORD 00473

Temp	Date	Time	s/ 210L
------	------	------	------------

VOID: RFI
12 06/14/23 00:30

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

Robinson #210266
Location



AIRSAS USA LLC (A-9)
2400
St. Louis, Mo. 63108
Ph: (314) 593-5100
Fax: (314) 593-7328

Certificate of Analysis

Test Date: 29-Nov-2021

Customer Name
Exclusive Supplier
Inoxirelens, Inc.
2081 Craig Road
St. Louis, Mo 63146

Lot # AG132803 Model 108

Exp. Date
24-Nov-2023

Cyl. Type
108
Component
Ehanol
Nitrogen
Certified Concentration
0.002 ± 0.002 BRAC (223 ppm)

Certification Traceable to N.I.S.T. RGM and to CRM Ehanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010581	103.6 ppm	EB0010562	104.2 ppm
EB0010581	52.12 ppm	EB0010579	52.81 ppm

CRM Serial No.	Concentration
CC4345688	800.0 ppm
CC234593	253.0 ppm

Analytical Method: NDIR

Permitted for use by the customer for the purpose of analysis
Reference: AOAC 2005.02 USA LLC (LMI)
Zhang 10/20/21 10/21

Approved for Release:

Rod Marsala
Rod Marsala

ISO 17025:2017 AZLA accredited. Certificate Number 3082.06
ISO 17034:2016 AZLA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

PERMIT TYPE II

WADE ROBINSON



I hereby authorize to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 308.111 through 308.119 RSMo.

DATE 11/18/2021
NUMBER 210266
EXPIRES 11/18/2023

Kevin A. Dwyer
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY
David A. Korman
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
Under No. 89

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The holder of this permit is authorized to operate the following breathalyzer(s):

Operator: **ROBINSON, WADE**
Permit No. **210266**
Date Issued **11/18/2021** Date Expires **11/18/2023**