



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| | | |
|---|---|----------------------------------|
| ALCO SENSOR IV SN 111634 | NAME OF AGENCY Missouri State Highway Patrol | DATE OF INSPECTION 07/02/2023 |
| LOCATION OF INSTRUMENT (STREET AND CITY) Troop B Zone 5 Laharpe Street in Kirksville | | TIME OF INSPECTION 6:17 pm |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

| | |
|--|---|
| <input checked="" type="checkbox"/> SIMULATOR SOLUTION | <input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE |
| <input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Guth Laboratories</u> LOT # <u>22430</u> EXP. DATE <u>11/30/2024</u> | |
| <input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>34.00</u> SIM. SN <u>MP2519</u> SIM. NIST EXP DATE <u>05/02/2024</u> | |

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

| | | |
|---------------|---------------|---------------|
| TEST 1 0.099 | TEST 2 0.097 | TEST 3 0.097 |
|---------------|---------------|---------------|

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

| | | | | | | | | | | | |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|
| REFUSALS | 0 | (0-.04) | 0 | (.05-.09) | 0 | (.10-.14) | 0 | (.15-.19) | 1 | (OVER .19) | 1 |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
Time Corrected

| | |
|--|------------------------------------|
| INSPECTING OFFICER | |
| SIGNATURE | PRINT NAME Riley D. Johnson |
| TYPE II PERMIT NUMBER/EXPIRATION DATE 220044// 02/10/2024 | TELEPHONE NUMBER (660) 385-2132 |

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.



GUTH LABORATORIES, INC.

890 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-684-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 22430 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on December 1, 2022, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1216% (w/vol) ethyl alcohol. The expiration date for this lot number is November 30, 2024 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

JUL 02 2023

AS IV Serial no: 111634
Version no: 532B

TEST RECORD 00281

Temp Date Time ^{s/} 210L

Air Blank:
07/02/23 18:23 .000
Calibration Check:
23 07/02/23 18:23 .099

Subject Name

Test #1

Subject I.D.

220044

Operator Name, I.D.

R.D. Johnson

Location

Troop B Zone 5

AS IV Serial no: 111634
Version no: 532B

TEST RECORD 00282

Temp Date Time ^{s/} 210L

Air Blank:
07/02/23 18:25 .000
Calibration Check:
24 07/02/23 18:25 .097

Subject Name

Test #2

Subject I.D.

220044

Operator Name, I.D.

R.D. Johnson

Location

Troop B Zone 5

AS IV Serial no: 111634
Version no: 532B

TEST RECORD 00283

Temp Date Time ^{s/} 210L

Air Blank:
07/02/23 18:26 .000
Calibration Check:
25 07/02/23 18:26 .097

Subject Name

Test #3

Subject I.D.

220044

Operator Name, I.D.

R.D. Johnson

Location

Troop B Zone 5

AS IV Serial no: 000000
Version no: 532B

TEST RECORD 00285

Temp Date Time ^{s/} 210L

VOID: RFI
12 07/02/23 18:29

Subject Name

Test - RFI

Subject I.D.

220044

Operator Name, I.D.

R.D. Johnson

Location

Troop B Zone 5



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
RILEY D. JOHNSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/10/2022

NUMBER 220044

EXPIRES 2/10/2024

Laura Q. Way

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Riley D. Johnson

, acting director
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator JOHNSON, RILEY
 Permit No 220044
 Date Issued 2/10/2022 Date Expires 2/10/2024

