



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

By Tracy Crews at 9:11 am, Sep 12, 2023

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111634	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 09/07/2023
-----------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 2920 N. Shamrock Rd, Jefferson City	TIME OF INSPECTION 8:40 am
---	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Guth</u>	LOT # <u>23180</u> EXP. DATE <u>05/17/2025</u>
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>33.99</u>	SIM. SN <u>MP2316</u> SIM. NIST EXP DATE <u>10/07/2023</u>

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .100	TEST 2 .100	TEST 3 .103
--------------	--------------	--------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	1	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Factory services performed. Memory re-initialized and reloaded calibrations and reprogrammed serial number. Replaced Mech diaphragm and Mech retainers. New 9volt battery installed. Instrument calibrated.

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Jimmy L. Cleveland
TYPE II PERMIT NUMBER/EXPIRATION DATE 230082 / 050325	TELEPHONE NUMBER (573) 751-4722

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111634
Version no: 532B

TEST RECORD 00026

Temp	Date	Time	s/ 210L
Air Blank:			
	09/07/23	08:44	.000
Calibration:			
	23 09/07/23	08:44	.100

Air Blank:
09/07/23 08:44 .000
Calibration:
23 09/07/23 08:44 .100

Subject Name

Cal
Subject I.D.

Operator Name, I.D.

Jimmy Cleveland

Location

2920 N. Sharnock Rd

Jefferson City

AS IV Serial no: 111634
Version no: 532B

TEST RECORD 00027

Temp	Date	Time	s/ 210L
Air Blank:			
	09/07/23	08:49	.000
Calibration Check:			
	24 09/07/23	08:49	.100

Air Blank:
09/07/23 08:49 .000
Calibration Check:
24 09/07/23 08:49 .100

Subject Name

Acc 1
Subject I.D.

Operator Name, I.D.

Jimmy L. Cleveland

Location

2920 N. Sharnock Rd

Jefferson City

AS IV Serial no: 111634
Version no: 532B

TEST RECORD 00028

Temp	Date	Time	s/ 210L
Air Blank:			
	09/07/23	08:54	.000
Calibration Check:			
	25 09/07/23	08:54	.100

Air Blank:
09/07/23 08:54 .000
Calibration Check:
25 09/07/23 08:54 .100

Subject Name

Acc 2
Subject I.D.

Operator Name, I.D.

Jimmy L. Cleveland

Location

2920 N. Sharnock Rd

Jefferson City

AS IV Serial no: 111634
Version no: 532B

TEST RECORD 00029

Temp	Date	Time	s/ 210L
Air Blank:			
	09/07/23	09:04	.000
Calibration Check:			
	25 09/07/23	09:04	.103

Air Blank:
09/07/23 09:04 .000
Calibration Check:
25 09/07/23 09:04 .103

Subject Name

Acc 3
Subject I.D.

Operator Name, I.D.

Jimmy L. Cleveland

Location

2920 N. Sharnock Rd

Jefferson City

AS IV Serial no: 111634
Version no: 532B

TEST RECORD 00030

Temp	Date	Time	s/ 210L
VOID: RFI			
	12 09/07/23	09:06	

VOID: RFI
12 09/07/23 09:06

Subject Name

RFI
Subject I.D.

Operator Name, I.D.

Jimmy L. Cleveland

Location

2920 N. Sharnock Rd

Jefferson City



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23180** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1220%** (w/vol) ethyl alcohol. The expiration date for this lot number is **May 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN11172002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

JIMMY L. CLEVELAND

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repair and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/3/2023

NUMBER 230082

EXPIRES 5/3/2025

Mike Massman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R)

 STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator CLEVELAND, JIMMY
Permit No 230082
Date issued 5/3/2023 Date Expires 5/3/2025

