

AS IV Serial no: 111657
Version no: 532B

TEST RECORD 00521

TEMP	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
12/12/23 11:12 .000
Calibration Check:
24 12/12/23 11:12 .103

Subject Name

Maint Test #1

Subject I.D.

Operator Name, I.D.

Cpl. W.J. Wzenko (220184)

Location

Troop A HQ

AS IV Serial no: 111657
Version no: 532B

TEST RECORD 00522

TEMP	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
12/12/23 11:13 .000
Calibration Check:
25 12/12/23 11:13 .103

Subject Name

Maint. Test #2

Subject I.D.

Operator Name, I.D.

Cpl. W.J. Wzenko (220184)

Location

Troop A HQ

AS IV Serial no: 111657
Version no: 532B

TEST RECORD 00523

TEMP	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
12/12/23 11:15 .000
Calibration Check:
25 12/12/23 11:15 .104

Subject Name

Maint. Test #3

Subject I.D.

Operator Name, I.D.

Cpl. W.J. Wzenko (220184)

Location

Troop A HQ

AS IV Serial no: 111657
Version no: 532B

TEST RECORD 00524

TEMP	Date	Time	s/ 210L
------	------	------	------------

VOID: RFI
12 12/12/23 11:17

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

Cpl. W.J. Wzenko (220184)

Location

Troop A HQ



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23180** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1220%** (w/vol) ethyl alcohol. The expiration date for this lot number is **May 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN11172002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

WALTER J. LUZENKO

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/22/2022

NUMBER 220184

EXPIRES 7/22/2024

Mike M...

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580 (7/71) (6-15)

LAB 4 (R5-15)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator LUZENKO, WALTER
 Permit No 220184
 Date Issued 7/22/2022 Date Expires 7/22/2024

