



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111659	PRINTER SN 09B.3589.478	DATE OF INSPECTION 01/10/2023
LOCATION OF INSTRUMENT (STREET AND CITY) 4001 NE Lakewood Way Ct., Lee's Summit, Mo 64064		TIME OF INSPECTION 8:09 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters</u> LOT # <u>AG114701</u> EXP. DATE <u>05/27/2023</u>	
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____	
<input checked="" type="checkbox"/> CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT) Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)	
<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE <input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE <input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE	

TEST 1 ← .098	TEST 2 ← .097	TEST 3 ← .097
---------------	---------------	---------------

- RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	2	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	2	(OVER .19)	2
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).  
 Instrument meets all DOHSS standards and guidelines.

INSPECTING OFFICER	
SIGNATURE <i>[Signature]</i>	PRINT NAME Dep. S. Stoff #63/0411
TYPE II PERMIT NUMBER/EXPIRATION DATE 220266 12/02/2024	TELEPHONE NUMBER (816) 524-4302

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901

AS IV Serial no: 111659  
Version no: 532B  
TEST RECORD 00828 g/  
Temp Date Time 210L  
VOID: RFI  
12 01/10/23 20:13  
Subject Name  
Monthly Maint  
Subject I.D.  
TEST RFD  
Operator Name, I.D.  
STAFF 603  
Location  
B/CSD 410

AS IV Serial no: 111659  
Version no: 532B  
TEST RECORD 00827 g/  
Temp Date Time 210L  
Air Blank: 01/10/23 20:12 .000  
Calibration Check: 24 01/10/23 20:12 .097  
Subject Name  
Monthly Maint  
Subject I.D.  
TEST 3  
Operator Name, I.D.  
STAFF 603  
Location  
6710

AS IV Serial no: 111659  
Version no: 532B  
TEST RECORD 00826 g/  
Temp Date Time 210L  
Air Blank: 01/10/23 20:10 .000  
Calibration Check: 24 01/10/23 20:10 .097  
Subject Name  
Monthly Maint  
Subject I.D.  
TEST #2  
Operator Name, I.D.  
STAFF 43  
Location  
6710

AS IV Serial no: 111659  
Version no: 532B  
TEST RECORD 00825 g/  
Temp Date Time 210L  
Air Blank: 01/10/23 20:09 .000  
Calibration Check: 23 01/10/23 20:09 .098  
Subject Name  
Monthly Maint  
Subject I.D.  
TEST 1  
Operator Name, I.D.  
STAFF 603  
Location  
6710



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**

*Exclusive Supplier*  
Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

**Test Date:** 1-Jun-2021

**Lot #** AG114701 **Model** 108cacc

<b>Exp. Date</b>	<b>Cyl. Type</b>	<b>Component</b>	<b>Certified Concentration</b>
27-May-2023	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

**Analytical Method:** NDIR

Digitally signed by Quality Control  
Date: 2021.06.03 17:37:33 -05:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

**ISO 17025:2017 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**SEAN R. STOFF**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/2/2022

NUMBER 220266

EXPIRES 12/2/2024

*Mike Mason*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Paula J. Nielson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator STOFF, SEAN  
 Permit No 220266  
 Date Issued 12/2/2022 Date Expires 12/2/2024

