

AS IV Serial no: 111663
Version no: 532B

TEST RECORD 00155

Temp Date Time 210L ^{g/}

Air Blank: 06/26/23 18:04 .000

Calibration Check: 28 06/26/23 18:04 .101

Subject Name

Subject I.D.

Operator Name, I.D.

Z.L. Ricker 220251

Location

34 Carter 104

Van Buren

AS IV Serial no: 111663
Version no: 532B

TEST RECORD 00156

Temp Date Time 210L ^{g/}

Air Blank: 06/26/23 18:05 .000

Calibration Check: 28 06/26/23 18:05 .100

Subject Name

Subject I.D.

Operator Name, I.D.

Z.L. Ricker 220251

Location

34 Carter 104

Van Buren

AS IV Serial no: 111663
Version no: 532B

TEST RECORD 00157

Temp Date Time 210L ^{g/}

Air Blank: 06/26/23 18:07 .000

Calibration Check: 29 06/26/23 18:07 .100

Subject Name

Subject I.D.

Operator Name, I.D.

Z.L. Ricker 220251

Location

34 Carter 104

Van Buren

AS IV Serial no: 111663
Version no: 532B

TEST RECORD 00158

Temp Date Time 210L ^{g/}

U01D: RFI

12 06/26/23 18:09

Subject Name

Subject I.D.

Operator Name, I.D.

Z.L. Ricker

Location

34 Carter 104

Van Buren



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **22430** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 1, 2022**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1216%** (w/vol) ethyl alcohol. The expiration date for this lot number is **November 30, 2024** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
ZACHARY RICKER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/21/2022

NUMBER 220251

EXPIRES 10/21/2024

Mike Masom

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator RICKER, ZACHARY
Permit No 220251
Date issued 10/21/2022 Date Expires 10/21/2024

