



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**

By Tracy Crews at 8:36 am, Oct 05, 2023

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111677	PRINTER SN 09B.3589.544	DATE OF INSPECTION 10/04/2023
LOCATION OF INSTRUMENT (STREET AND CITY) Madison County Jail (124 N. Main, Fredericktown, MO 63645)		TIME OF INSPECTION 9:55 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories LOT # 23180 EXP. DATE 05/17/2025

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN MP2207 SIMULATOR EXP DATE 03/10/2024

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .099

TEST 2 .099

TEST 3 .099

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	65	(.05-.09)	1	(.10-.14)	0	(.15-.19)	1	(OVER .19)	0
----------	---	---------	----	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Advanced time forward one minute.

**INSPECTING OFFICER**

SIGNATURE

PRINT NAME  
E. J. Ganime

TYPE II PERMIT NUMBER/EXPIRATION DATE  
210281 12/7/2023

TELEPHONE NUMBER  
(573) 840-9500

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901

AS IV Serial no: 111577  
Version no: 532B

TEST RECORD 01091

Temp Date Time 210L

Air Blank:  
10/04/23 10:12 .000  
Calibration Check:  
22 10/04/23 10:12 .099

Subject Name

TEST 1

Subject I.D.

Operator Name, I.D.

E.J. GANIME 210281

Location

MADISON CO. IAZL

AS IV Serial no: 111577  
Version no: 532B

TEST RECORD 01092

Temp Date Time 210L

Air Blank:  
10/04/23 10:13 .000  
Calibration Check:  
22 10/04/23 10:13 .099

Subject Name

TEST 2

Subject I.D.

Operator Name, I.D.

E.J. GANIME 210281

Location

MADISON CO. IAZL

AS IV Serial no: 111577  
Version no: 532B

TEST RECORD 01093

Temp Date Time 210L

Air Blank:  
10/04/23 10:15 .000  
Calibration Check:  
22 10/04/23 10:15 .099

Subject Name

TEST 3

Subject I.D.

Operator Name, I.D.

E.J. GANIME 210281

Location

MADISON CO. IAZL

AS IV Serial no: 111577  
Version no: 532B

TEST RECORD 01094

Temp Date Time 210L

UOID: RFI  
12 10/04/23 10:16

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

E.J. GANIME 210281

Location

MADISON CO. IAZL



## GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23180** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1220%** (w/vol) ethyl alcohol. The expiration date for this lot number is **May 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN11172002 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**ERIC J. GANIME**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/7/2021

NUMBER 210281

EXPIRES 12/7/2023

*Laura A. Way*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Donald B. Krumm*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator GANIME, ERIC  
Permit No 210281  
Date Issued 12/7/2021 Date Expires 12/7/2023

