

RECEIVED

By Tracy Crews at 7:29 am, May 10, 2023



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| | | |
|---|---|----------------------------------|
| ALCO SENSOR IV SN 111744 | NAME OF AGENCY BOONE COUNTY SHERIFF'S OFFICE | DATE OF INSPECTION 05/09/2023 |
| LOCATION OF INSTRUMENT (STREET AND CITY) 2111 E. COUNTY DR, COLUMBIA | | TIME OF INSPECTION 9:26 pm |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER INTOXIMETERS LOT # AG231902 EXP. DATE 11/15/2024

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .101

TEST 2 ← .100

TEST 3 ← .100

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

| | | | | | | | | | | | |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|
| REFUSALS | 1 | (0-.04) | 0 | (.05-.09) | 0 | (.10-.14) | 1 | (.15-.19) | 0 | (OVER .19) | 1 |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Operating within standard.

INSPECTING OFFICER

SIGNATURE

#6067

PRINT NAME

Jesse Dennison

TYPE II PERMIT NUMBER/EXPIRATION DATE

220166, 6/24/2024

TELEPHONE NUMBER

(573) 875-1111

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: J11744
Version no: 532B

TEST RECORD 00710

Temp Date Time 210L
Air Blank: 05/09/23 21:26 .000
Calibration Check: 23 05/09/23 21:26 .100

Subject Name

Test 2

Subject I, J.

Operator Name, J, J.

Demison, 22066

Location

211 County Dr

AS IV Serial no: J11744
Version no: 532B

TEST RECORD 00711

Temp Date Time 210L
Air Blank: 05/09/23 21:30 .000
Calibration Check: 24 05/09/23 21:30 .100

Subject Name

Test 3

Subject I, J.

Operator Name, J, J.

Demison, 22066

Location

211 County Dr

AS IV Serial no: J11744
Version no: 532B

TEST RECORD 00712

Temp Date Time 210L
Air Blank: 05/09/23 21:32 .000
Calibration Check: 25 05/09/23 21:32 .100

Subject Name

RFI

Subject I, J.

Operator Name, J, J.

Demison, 22066

Location

211 County Dr



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 17-Nov-2022

Lot # AG231902 **Model** 108

| | | | |
|--------------------------------|-------------------------|---|---|
| Exp Date 15-Nov-2024 | Cyl. Type 108 | Component Ethanol Nitrogen | Certified Concentration 0.100 ± 2% BrAC (272 ppm) |
|--------------------------------|-------------------------|---|---|

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

| RGM Serial No. | Concentration | RGM Serial No. | Concentration |
|----------------|---------------|----------------|---------------|
| EB0010581 | 391.8 ppm | EB0010603 | 392.5 ppm |
| EB0010570 | 259.8 ppm | EB0010559 | 258.9 ppm |
| EB0010285 | 209.0 ppm | EB0010562 | 104.2 ppm |
| EB0010561 | 103.7 ppm | EB0010579 | 52.94 ppm |
| EB0010681 | 52.22 ppm | | |

| CRM Serial No. | Concentration | CRM Serial No. | Concentration |
|----------------|---------------|----------------|---------------|
| CC727481 | 800.0 ppm | CC727493 | 390.0 ppm |
| CC727496 | 253.0 ppm | CC727498 | 150.0 ppm |

Analytical Method: NDIR

Digitally signed by: Quality Control
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)
 Date: 11.17.2022 20:17

Approved for Release: _____

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
JESSE DENNISON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/24/2022

Mike Massum

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 220166

David J. Nickelson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 6/24/2024

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator DENNISON, JESSE
 Permit No 220166
 Date Issued 6/24/2022 Date Expires 6/24/2024

