

AS IU Serial no: 111748
Version no: 532B

TEST RECORD 00136
Temp Date Time 210L
s/

Air Blank:
04/02/23 18:20 .000
Calibration Check:
22 04/02/23 18:20 .098

Subject Name
Acc. #1
Subject I.D.

Operator Name, I.D.

T. WRIGHT, 1996

Location

1810 SELMORE BLVD

DRIVE

AS IU Serial no: 111748
Version no: 532B

TEST RECORD 00137
Temp Date Time 210L
s/

Air Blank:
04/02/23 18:24 .000
Calibration Check:
22 04/02/23 18:24 .097

Subject Name
Acc. #2
Subject I.D.

Operator Name, I.D.

T. WRIGHT, 1996

Location

1810 SELMORE BLVD

DRIVE

AS IU Serial no: 111748
Version no: 532B

TEST RECORD 00138
Temp Date Time 210L
s/

Air Blank:
04/02/23 18:27 .000
Calibration Check:
23 04/02/23 18:27 .097

Subject Name
Acc. #3
Subject I.D.

Operator Name, I.D.

T. WRIGHT, 1996

Location

1810 SELMORE BLVD

DRIVE

AS IU Serial no: 111748
Version no: 532B

TEST RECORD 00139
Temp Date Time 210L
s/

VOID: RT1
12 04/02/23 18:29

Subject Name
RF1 TEST
Subject I.D.

Operator Name, I.D.

T. WRIGHT, 1996

Location

1810 SELMORE BLVD

DRIVE

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 21001
EXPIRATION DATE: June 16, 2023 at 11:59 p.m.

RepCo Marketing Co. certifies the following:

RepCo Marketing Co. prepared, tested and supplied Lot Number 21001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1222 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 17, 2021 The expiration date for this lot number is June 16, 2023 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
TYLER WRIGHT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/14/2022

NUMBER 220161

EXPIRES 6/14/2024

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nickelson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator WRIGHT, TYLER
Permit No 220161
Date Issued 6/14/2022 **Date Expires** 6/14/2024

