

AS IV Serial no: 111755
Version no: 532B

TEST RECORD 00504 ^{s/}

Temp Date Time 210L

Air Blank: 04/19/23 14:55 .000

Calibration Check: 25 04/19/23 14:55 .100

Subject Name

Subject I.D.

Operator Name, I.D.

I. MEYER 220218

Location

ZONE 2 OFFICE

HOWELL COUNTY

AS IV Serial no: 111755
Version no: 532B

TEST RECORD 00505 ^{s/}

Temp Date Time 210L

Air Blank: 04/19/23 14:57 .000

Calibration Check: 25 04/19/23 14:57 .099

Subject Name

Subject I.D.

Operator Name, I.D.

I. MEYER 220218

Location

ZONE 7 OFFICE

HOWELL COUNTY

AS IV Serial no: 111755
Version no: 532B

TEST RECORD 00506 ^{s/}

Temp Date Time 210L

Air Blank: 04/19/23 14:59 .000

Calibration Check: 26 04/19/23 14:59 .099

Subject Name

Subject I.D.

Operator Name, I.D.

I. MEYER 220218

Location

HOWELL COUNTY

ZONE 7 OFFICE

AS IV Serial no: 111755
Version no: 532B

TEST RECORD 00507 ^{s/}

Temp Date Time 210L

VOID: RFI

12 04/19/23 15:00

Subject Name

Subject I.D.

Operator Name, I.D.

I. MEYER 220218

Location

ZONE 7 OFFICE

HOWELL COUNTY

RepCo Marketing Co
3101-188 Stony Brook Drive
Raleigh, NC 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 21001
EXPIRATION DATE: June 16, 2023 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 21001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1222 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 17, 2021 The expiration date for this lot number is June 16, 2023 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.


Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
THOMAS W. MEYER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/7/2022

NUMBER 220218

EXPIRES 9/7/2024

Mike Morrison

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Dave J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-19)

LAB 4 (R5 15)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MEYER, THOMAS
Permit No 220218
Date issued 9/7/2022 **Date Expires** 9/7/2024

