

RECEIVED
By Tracy Crews at 7:32 am, Dec 18, 2023



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111760	NAME OF AGENCY JOHNSON COUNTY SHERIFF'S OFFICE	DATE OF INSPECTION 12/14/2023
LOCATION OF INSTRUMENT (STREET AND CITY) 278 SOUTH WEST 871 ROAD CENTERVIEW, MISSOURI 64019-8300		TIME OF INSPECTION 10:21 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER INTOXIMETERS LOT # AG310903 EXP. DATE 04/19/2025
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIM. SN _____ SIM. NIST EXP. DATE _____

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 → .099	TEST 2 → .099	TEST 3 → .098
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME PATRICK W. BOYER
TYPE II PERMIT NUMBER/EXPIRATION DATE 230100 / 05-30-2025	TELEPHONE NUMBER (660) 747-6469

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111760
Version no: 532B

TEST RECORD 00733

Temp Date Time ^{s/} 210L

Air Blank: 12/14/23 22:21 .000
Calibration Check: 22 12/14/23 22:21 .099

Subject Name
TEST #1

Subject I.D.

Operator Name, I.D.

Boyer 230100

Location

JCSO

AS IV Serial no: 111760
Version no: 532B

TEST RECORD 00734

Temp Date Time ^{s/} 210L

Air Blank: 12/14/23 22:23 .000
Calibration Check: 22 12/14/23 22:23 .099

Subject Name

TEST #2

Subject I.D.

Operator Name, I.D.

Boyer 230100

Location

JCSO

AS IV Serial no: 111760
Version no: 532B

TEST RECORD 00735

Temp Date Time ^{s/} 210L

Air Blank: 12/14/23 22:24 .000
Calibration Check: 23 12/14/23 22:24 .098

Subject Name

TEST #3

Subject I.D.

Operator Name, I.D.

Boyer 230100

Location

JCSO

AS IV Serial no: 111760
Version no: 532B

TEST RECORD 00736

Temp Date Time ^{s/} 210L

VOID: RFI
12 12/14/23 22:25

Subject Name

TEST RFI

Subject I.D.

Operator Name, I.D.

Boyer 230100

Location

JCSO

AS IV Serial no: 111760
Version no: 532B

TEST RECORD 00737

Temp Date Time ^{s/} 210L

Air Blank: 12/14/23 22:27 .000
Subject Test: Auto
23 12/14/23 22:27 .000

Subject Name

SCIF TEST

Subject I.D.

Operator Name, I.D.

Boyer 230100

Location

JCSO



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
PATRICK W. BOYER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/30/2023

NUMBER 230100

EXPIRES 5/30/2025

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Douglas J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0721 (8-10)

LAB-4 (R6-10)

 **STATE OF MISSOURI**
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **BOYER, PATRICK**
Permit No **230100**
Date Issued **5/30/2023** Date Expires **5/30/2025**





Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 20-Apr-2023

Lot # AG310903 **Model** 108

Exp Date	Cyl. Type	Component	Certified Concentration
19-Apr-2025	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm)

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by:Quality Control
Reason:Dry gas standard certification of analysis
Location:Airgas USA LLC (Lab)
Date:04.20.2023 15:24

Approved for Release:

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07