



AS IV Serial no: 111762  
Version no: 532B

TEST RECORD 00738

| Temp | Date | Time | s/ | 210L |
|------|------|------|----|------|
|------|------|------|----|------|

Air Blank:  
08/04/23 06:14 .000  
Calibration Check:  
19 08/04/23 06:14 .000

Subject Name  
Blank Check  
Subject I.D.

Operator Name, I.D.

K. Bailey 230099

Location

Stone Co Jail

AS IV Serial no: 111762  
Version no: 532B

TEST RECORD 00741

| Temp | Date | Time | s/ | 210L |
|------|------|------|----|------|
|------|------|------|----|------|

Air Blank:  
08/04/23 06:27 .000  
Calibration Check:  
21 08/04/23 06:27 .103

Subject Name  
Test 3  
Subject I.D.

Operator Name, I.D.

Bailey 230099

Location

Stone Co Jail

AS IV Serial no: 111762  
Version no: 532B

TEST RECORD 00739

| Temp | Date | Time | s/ | 210L |
|------|------|------|----|------|
|------|------|------|----|------|

Air Blank:  
08/04/23 06:19 .000  
Calibration Check:  
20 08/04/23 06:19 .102

Subject Name  
Test 1  
Subject I.D.

Operator Name, I.D.

Bailey 230099

Location

Stone Co Jail

AS IV Serial no: 111762  
Version no: 532B

TEST RECORD 00740

| Temp | Date | Time | s/ | 210L |
|------|------|------|----|------|
|------|------|------|----|------|

Air Blank:  
08/04/23 06:23 .000  
Calibration Check:  
21 08/04/23 06:23 .103

Subject Name  
Test 2  
Subject I.D.

Operator Name, I.D.

Bailey 230099

Location

Stone Co Jail

AS IV Serial no: 111762  
Version no: 532B

TEST RECORD 00742

| Temp | Date | Time | s/ | 210L |
|------|------|------|----|------|
|------|------|------|----|------|

VOID: RFI  
12 08/04/23 06:29

Subject Name  
RFI  
Subject I.D.

Operator Name, I.D.

Bailey 230099

Location

Stone Co Jail



## GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **22430** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 1, 2022**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1216%** (w/vol) ethyl alcohol. The expiration date for this lot number is **November 30, 2024** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L  $\pm$  3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*





STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**KELLYN BAILEY**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/30/2023

NUMBER 230099

EXPIRES 5/30/2025

*Mike Mason*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*David J. Nielson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator BAILEY, KELLYN  
 Permit No 230099  
 Date Issued 5/30/2023 Date Expires 5/30/2025

