



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111773	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 09/16/2023
-----------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Zone 4/15 Office - 275 B Southwest Outer Road, Branson	TIME OF INSPECTION 8:47 pm
--	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories LOT # 23180 EXP. DATE 05/17/2025

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 33.99 SIM. SN MP2422 SIM. NIST EXP DATE 02/22/2024

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .099

TEST 2 .100

TEST 3 .099

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Calibrated and changed printer ribbon.

INSPECTING OFFICER

SIGNATURE

PRINT NAME

Brandon Hall

TYPE II PERMIT NUMBER/EXPIRATION DATE

230011 / 01/24/2025

TELEPHONE NUMBER

(417) 895-6868

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IU Serial no: 111773
Version no: 532B

TEST RECORD 00154

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank	08/16/23	20:47	.000	
Calibration Check:	25	08/16/23	20:47	.099

Subject Name

Acc check

Subject I.D.

Brandon Hall 347

Operator Name, I.D.

Zone 4/15 office

Location

Branson

AS IU Serial no: 111773
Version no: 532B

TEST RECORD 00156

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank	08/16/23	20:53	.000	
Calibration Check:	24	08/16/23	20:53	.099

Subject Name

Acc check

Subject I.D.

Brandon Hall 347

Operator Name, I.D.

Zone 4/15 office

Location

Branson

AS IU Serial no: 111773
Version no: 532B

TEST RECORD 00155

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank	08/16/23	20:50	.000	
Calibration Check:	26	08/16/23	20:50	.100

Subject Name

Acc check

Subject I.D.

Brandon Hall 347

Operator Name, I.D.

Zone 4/15 office

Location

Branson

AS IU Serial no: 111773
Version no: 532B

TEST RECORD 00157

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank	08/16/23	20:55	.000	
-----------	----------	-------	------	--

Subject Name

Acc check

Subject I.D.

Brandon Hall 347

Operator Name, I.D.

Zone 4/15 office

Location

Branson

08/16/23 20:43 .000
Calibration
23 08/16/23 20:43 .100

Subject Name

Acc Check

Subject I.D.

Brandon Hall 347

Operator Name, I.D.

Zone 4/15 office

Location

Branson



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23180** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1220%** (w/vol) ethyl alcohol. The expiration date for this lot number is **May 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN11172002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
BRANDON A. HALL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/24/2023

NUMBER 230011

EXPIRES 1/24/2025

Mike Massman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nielson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HALL, BRANDON
Permit No 230011
Date Issued 1/24/2023 **Date Expires** 1/24/2025

