



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

|  |   |                                  |
|--|---|----------------------------------|
| ALCO SENSOR IV SN<br>111774  | NAME OF AGENCY<br>Carroll County Sheriff's Office | DATE OF INSPECTION<br>03/04/2023 |
| LOCATION OF INSTRUMENT (STREET AND CITY)<br>106 S. Folger St, Carrollton |   | TIME OF INSPECTION<br>2:49 pm    |

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Guth Labs LOT # 22430 EXP. DATE 11/30/2024
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.02 SIM. SN MP3879 SIM. NIST EXP DATE 05/02/2023

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

|               |               |               |
|---------------|---------------|---------------|
| TEST 1 ← .104 | TEST 2 ← .104 | TEST 3 ← .103 |
|---------------|---------------|---------------|

- RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

|          |   |         |   |           |   |           |   |           |   |            |   |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|
| REFUSALS | 0 | (0-.04) | 0 | (.05-.09) | 0 | (.10-.14) | 0 | (.15-.19) | 0 | (OVER .19) | 0 |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

|  |                                    |
|--|------------------------------------|
| SIGNATURE<br><i>Brian C. Woods</i>                         | PRINT NAME<br>Brian C. Woods       |
| TYPE II PERMIT NUMBER/EXPIRATION DATE<br>220181 07/12/2024 | TELEPHONE NUMBER<br>(660) 542-2200 |

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111774  
Version no: 5328

TEST RECORD 00407

Temp Date Time <sup>a/</sup> 210L

Air Blank  
03/04/23 14:49 .000  
Calibration Check:  
18 03/04/23 14:49 .104

Subject Name  
Test #1  
Subject I.D.

Operator Name, I.D.  
B.C. Woods, 782  
Location

AS IV Serial no: 111774  
Version no: 5328

TEST RECORD 00408

Temp Date Time <sup>a/</sup> 210L

Air Blank  
03/04/23 14:51 .000  
Calibration Check:  
19 03/04/23 14:51 .104

Subject Name  
Test #2  
Subject I.D.

Operator Name, I.D.  
B.C. Woods, 782  
Location

AS IV Serial no: 111774  
Version no: 5328

TEST RECORD 00411

Temp Date Time <sup>a/</sup> 210L

Air Blank  
03/04/23 14:52 .000  
Calibration Check:  
20 03/04/23 14:52 .103

Subject Name  
Test #3  
Subject I.D.

Operator Name, I.D.  
B.C. Woods, 782  
Location

AS IV Serial no: 111774  
Version no: 5328

TEST RECORD 00410

Temp Date Time <sup>a/</sup> 210L

VOID: RFI  
12 03/04/23 14:05

Subject Name  
RFI Test  
Subject I.D.

Operator Name, I.D.  
B.C. Woods, 782  
Location

AS IV Serial no: 111774  
Version no: 5328

TEST RECORD 00411

Temp Date Time <sup>a/</sup> 210L

Air Blank  
03/04/23 14:06 .000  
Subject Test: Auto  
22 03/04/23 14:06 .000

Subject Name  
Blank Test  
Subject I.D.

Operator Name, I.D.  
B.C. Woods, 782  
Location



## GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4811 • TELEPHONE: 717-664-6470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **22430** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 1, 2022**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1216%** (w/vol) ethyl alcohol. The expiration date for this lot number is **November 30, 2024** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L**  $\pm 3\%$ .

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**BRIAN C. WOODS**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/12/2022

*Mike Masman*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 220181

*Paula J. Nielsen*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 7/12/2024

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator WOODS, BRIAN  
 Permit No 220181  
 Date Issued 7/12/2022 Date Expires 7/12/2024

